

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071618\
 Data File : VW003959.D
 Acq On : 16 Jul 2018 18:50
 Operator : VA/AP
 Sample : J3975-03
 Misc : 5.47G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-6

Quant Time: Jul 17 08:41:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	10043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	15080	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	11074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	4059	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	5717	50.93	ug/l	0.00
Spiked Amount			Recovery	=	101.86%	
35) Dibromofluoromethane	7.88	113	5183	53.57	ug/l	0.00
Spiked Amount			Recovery	=	107.14%	
50) Toluene-d8	10.33	98	16442	43.79	ug/l	0.00
Spiked Amount			Recovery	=	87.58%	
62) 4-Bromofluorobenzene	12.63	95	12167	89.87	ug/l	0.00
Spiked Amount			Recovery	=	179.74%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.13	43	6809	252.41	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071618\
Data File : VW003959.D
Acq On : 16 Jul 2018 18:50
Operator : VA/AP
Sample : J3975-03
Misc : 5.47G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-6

Quant Time: Jul 17 08:41:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

