

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071624\
 Data File : VW029588.D
 Acq On : 16 Jul 2024 13:31
 Operator : SY/MD
 Sample : VIBLK497
 Misc : 5.00g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK497

Quant Time: Jul 17 00:55:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 17 00:53:35 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	246361	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	165648	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	37251	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	53509	13.445	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.800%
7) Chloroethane-d5	2.893	69	45859	16.003	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.000%
11) 1,1-Dichloroethene-d2	4.021	65	24307	13.601	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.400%
21) 2-Butanone-d5	7.081	46	28128	23.544	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	47.080%
24) Chloroform-d	7.648	84	130534	18.328	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.320%
26) 1,2-Dichloroethane-d4	8.307	65	86279	20.654	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	8.276	84	230390	23.217	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.880%
36) 1,2-Dichloropropane-d6	9.270	67	83449	26.301	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.200%
41) Toluene-d8	10.318	98	165630	18.979	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.920%
43) trans-1,3-Dichloroprop...	10.581	79	23280	17.327	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.320%
47) 2-Hexanone-d5	10.922	63	17400	30.180	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	60.360%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	52687	20.453	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.800%
66) 1,2-Dichlorobenzene-d4	13.848	152	23617	18.205	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.800%#
Target Compounds						Qvalue
16) Methylene chloride	4.929	84	32197	7.300	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071624\
Data File : VW029588.D
Acq On : 16 Jul 2024 13:31
Operator : SY/MD
Sample : VIBLK497
Misc : 5.00g/10mL/MSVOA_W/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK497

Quant Time: Jul 17 00:55:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 17 00:53:35 2024
Response via : Initial Calibration

