

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071724\
 Data File : VW029612.D
 Acq On : 17 Jul 2024 13:32
 Operator : SY/MD
 Sample : P3263-16
 Misc : 5.06g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CK4

Quant Time: Jul 18 02:20:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 18 02:17:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/L	-8.84
28) Chlorobenzene-d5	0.000	117	0	0.000	ug/L	-11.63
58) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/L	-13.56
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.180	65	40	0.000	ug/L	-0.18
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.000%#
7) Chloroethane-d5	2.857	69	74	0.000	ug/L	-0.04
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.000%#
11) 1,1-Dichloroethene-d2	4.106	65	26	0.000	ug/L	0.09
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.000%#
21) 2-Butanone-d5	7.069	46	51	0.000	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
24) Chloroform-d	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	40 - 150	Recovery	=	0.000%#
26) 1,2-Dichloroethane-d4	8.087	65	48	0.000	ug/L	-0.22
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.000%#
32) Benzene-d6	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	20 - 135	Recovery	=	0.000%#
36) 1,2-Dichloropropane-d6	9.227	67	49	0.000	ug/L	-0.04
Spiked Amount	25.000	Range	70 - 120	Recovery	=	0.000%#
41) Toluene-d8	0.000	98	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 130	Recovery	=	0.000%#
43) trans-1,3-Dichloroprop...	10.666	79	42	0.000	ug/L	0.09
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.000%#
47) 2-Hexanone-d5	11.136	63	69	0.000	ug/L	0.21
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
56) 1,1,2,2-Tetrachloroeth...	12.635	84	42	0.000	ug/L	-0.05
Spiked Amount	25.000	Range	45 - 120	Recovery	=	0.000%#
66) 1,2-Dichlorobenzene-d4	0.000	152	0	0.000	ug/L	
Spiked Amount	25.000	Range	75 - 120	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071724\
Data File : VW029612.D
Acq On : 17 Jul 2024 13:32
Operator : SY/MD
Sample : P3263-16
Misc : 5.06g/10mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CK4

Quant Time: Jul 18 02:20:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jul 18 02:17:40 2024
Response via : Initial Calibration

