

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071724\
 Data File : VW029616.D
 Acq On : 17 Jul 2024 15:08
 Operator : SY/MD
 Sample : P3263-20
 Misc : 5.13g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CLO

Quant Time: Jul 18 02:21:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jul 18 02:17:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	366258	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	295123	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	91436	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	82889	14.009	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	56.040%		
7) Chloroethane-d5	2.899	69	67544	15.854	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	63.400%		
11) 1,1-Dichloroethene-d2	4.021	65	39582	14.898	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	59.600%		
21) 2-Butanone-d5	7.075	46	44142	24.853	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	49.700%		
24) Chloroform-d	7.648	84	177305	16.745	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	67.000%		
26) 1,2-Dichloroethane-d4	8.307	65	102638	16.527	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	66.120%#		
32) Benzene-d6	8.276	84	323610	18.304	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	73.200%		
36) 1,2-Dichloropropane-d6	9.270	67	109167	19.312	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	77.240%		
41) Toluene-d8	10.325	98	270913	17.424	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	69.680%		
43) trans-1,3-Dichloroprop...	10.575	79	29213	12.204	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	48.800%		
47) 2-Hexanone-d5	10.922	63	21766	21.190	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	42.380%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	68411	14.906	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	59.640%		
66) 1,2-Dichlorobenzene-d4	13.848	152	51512	16.177	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	64.720%#		
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	37840	5.771	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071724\
Data File : VW029616.D
Acq On : 17 Jul 2024 15:08
Operator : SY/MD
Sample : P3263-20
Misc : 5.13g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CLO

Quant Time: Jul 18 02:21:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jul 18 02:17:40 2024
Response via : Initial Calibration

