

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071725\
 Data File : VW031882.D
 Acq On : 17 Jul 2025 11:51
 Operator : SY/MD
 Sample : VW0717SBSD01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0717SBSD01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025

Quant Time: Jul 18 02:50:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	218368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	418079	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.636	117	367858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	170253	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	170444	54.388	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	= 108.780%	
35) Dibromofluoromethane	7.904	113	136361	49.983	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	= 99.960%	
50) Toluene-d8	10.325	98	524424	51.655	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	= 103.300%	
62) 4-Bromofluorobenzene	12.617	95	202301	54.147	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	= 108.300%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.052	85	32491	21.848	ug/l	100
3) Chloromethane	2.259	50	39583	22.077	ug/l	97
4) Vinyl Chloride	2.412	62	51040	21.658	ug/l	94
5) Bromomethane	2.826	94	37740	20.503	ug/l	99
6) Chloroethane	2.979	64	34249	21.395	ug/l	99
7) Trichlorofluoromethane	3.308	101	49277	23.025	ug/l	98
8) Diethyl Ether	3.722	74	35917	22.085	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.100	101	49413	20.801	ug/l	100
10) Methyl Iodide	4.308	142	73606	20.099	ug/l	99
11) Tert butyl alcohol	5.222	59	21932	112.964	ug/l	99
12) 1,1-Dichloroethene	4.076	96	54848	21.079	ug/l	95
13) Acrolein	3.936	56	38543	110.454	ug/l	92
14) Allyl chloride	4.710	41	84485	21.162	ug/l	96
15) Acrylonitrile	5.411	53	90150	112.726	ug/l	99
16) Acetone	4.161	43	68131	87.958	ug/l	96
17) Carbon Disulfide	4.423	76	136762	19.642	ug/l	96
18) Methyl Acetate	4.716	43	37149	16.781	ug/l	100
19) Methyl tert-butyl Ether	5.472	73	101726	23.034	ug/l	100
20) Methylene Chloride	4.960	84	65933	18.267	ug/l	95
21) trans-1,2-Dichloroethene	5.460	96	60123	21.746	ug/l	95
22) Diisopropyl ether	6.344	45	182967	23.499	ug/l	98
23) Vinyl Acetate	6.289	43	657811	123.690	ug/l	100
24) 1,1-Dichloroethane	6.252	63	111283	22.045	ug/l	99
25) 2-Butanone	7.197	43	117240	116.343	ug/l	99
26) 2,2-Dichloropropane	7.191	77	53272	18.088	ug/l	96
27) cis-1,2-Dichloroethene	7.191	96	70811	22.281	ug/l	99
28) Bromochloromethane	7.533	49	48138	21.610	ug/l	97
29) Tetrahydrofuran	7.557	42	81471	118.055	ug/l	99
30) Chloroform	7.697	83	119920	22.357	ug/l	95
31) Cyclohexane	7.972	56	97876	21.935	ug/l #	91
32) 1,1,1-Trichloroethane	7.892	97	84958	20.553	ug/l	99
36) 1,1-Dichloropropene	8.100	75	82489	20.895	ug/l	100
37) Ethyl Acetate	7.277	43	56207	22.570	ug/l	99
38) Carbon Tetrachloride	8.087	117	77394	19.238	ug/l	98
39) Methylcyclohexane	9.343	83	105322	21.443	ug/l	95
40) Benzene	8.337	78	249996	21.404	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071725\
 Data File : VW031882.D
 Acq On : 17 Jul 2025 11:51
 Operator : SY/MD
 Sample : VW0717SBSD01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0717SBSD01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025

Quant Time: Jul 18 02:50:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:35:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.502	41	31081	23.298	ug/l	99
42) 1,2-Dichloroethane	8.410	62	84445	21.397	ug/l	99
43) Isopropyl Acetate	8.441	43	95400	22.223	ug/l	97
44) Trichloroethene	9.105	130	59927	20.557	ug/l	98
45) 1,2-Dichloropropane	9.374	63	59816	21.190	ug/l	100
46) Dibromomethane	9.465	93	39615	21.737	ug/l	95
47) Bromodichloromethane	9.654	83	89815	21.016	ug/l	99
48) Methyl methacrylate	9.447	41	45191	21.409	ug/l	96
49) 1,4-Dioxane	9.471	88	9686	453.788	ug/l #	98
51) 4-Methyl-2-Pentanone	10.215	43	274460	111.209	ug/l	98
52) Toluene	10.392	92	160528	21.466	ug/l	96
53) t-1,3-Dichloropropene	10.611	75	85600	21.348	ug/l	98
54) cis-1,3-Dichloropropene	10.075	75	98325	21.631	ug/l	96
55) 1,1,2-Trichloroethane	10.788	97	51549	21.593	ug/l	97
56) Ethyl methacrylate	10.648	69	76651	23.667	ug/l	98
57) 1,3-Dichloropropane	10.934	76	91688	21.876	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.928	63	182347	102.247	ug/l	98
59) 2-Hexanone	10.971	43	194313	116.188	ug/l	100
60) Dibromochloromethane	11.136	129	56037	19.629	ug/l	97
61) 1,2-Dibromoethane	11.239	107	51000	21.340	ug/l	94
64) Tetrachloroethene	10.861	164	45287	18.821	ug/l #	89
65) Chlorobenzene	11.660	112	170081	20.956	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.727	131	54577	21.102	ug/l	98
67) Ethyl Benzene	11.733	91	303022	21.531	ug/l	97
68) m/p-Xylenes	11.837	106	231665	42.820	ug/l	100
69) o-Xylene	12.166	106	109483	21.856	ug/l	99
70) Styrene	12.178	104	191034	22.060	ug/l	99
71) Bromoform	12.349	173	30180	19.560	ug/l #	96
73) Isopropylbenzene	12.465	105	282352	21.803	ug/l	100
74) N-amyl acetate	12.270	43	90101	25.813	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.715	83	66690	23.196	ug/l	100
76) 1,2,3-Trichloropropane	12.769	75	52628m	23.920	ug/l	
77) Bromobenzene	12.745	156	62529	21.649	ug/l	96
78) n-propylbenzene	12.800	91	350510	22.602	ug/l	97
79) 2-Chlorotoluene	12.891	91	213392	22.851	ug/l	99
80) 1,3,5-Trimethylbenzene	12.940	105	240871	22.470	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.507	75	19627	20.778	ug/l	95
82) 4-Chlorotoluene	12.983	91	221327	22.575	ug/l	99
83) tert-Butylbenzene	13.202	119	196418	21.391	ug/l	97
84) 1,2,4-Trimethylbenzene	13.245	105	249572	22.978	ug/l	97
85) sec-Butylbenzene	13.379	105	310256	22.511	ug/l	99
86) p-Isopropyltoluene	13.495	119	251261	22.121	ug/l	99
87) 1,3-Dichlorobenzene	13.495	146	128627	22.206	ug/l	99
88) 1,4-Dichlorobenzene	13.574	146	127323	21.475	ug/l	95
89) n-Butylbenzene	13.818	91	246392	22.326	ug/l	99
90) Hexachloroethane	14.086	117	42404	20.693	ug/l	95
91) 1,2-Dichlorobenzene	13.867	146	117158	22.129	ug/l	94
92) 1,2-Dibromo-3-Chloropr...	14.483	75	12398	22.085	ug/l	96
93) 1,2,4-Trichlorobenzene	15.129	180	68381	20.943	ug/l	99
94) Hexachlorobutadiene	15.226	225	30297	19.216	ug/l	90
95) Naphthalene	15.360	128	189006	23.612	ug/l	99
96) 1,2,3-Trichlorobenzene	15.549	180	59707	19.901	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071725\
Data File : VW031882.D
Acq On : 17 Jul 2025 11:51
Operator : SY/MD
Sample : VW0717SBSD01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025

Quant Time: Jul 18 02:50:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:35:17 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.965 min Scan# 99
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion:168 Resp: 21836
Ion Ratio Lower Upper
168 100
99 68.7 49.4 74.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#2
Dichlorodifluoromethane
Concen: 21.848 ug/l
RT: 2.052 min Scan# 27
Delta R.T. 0.013 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion: 85 Resp: 32491
Ion Ratio Lower Upper
85 100
87 30.9 15.4 46.2









#3
Chloromethane
Concen: 22.077 ug/l
RT: 2.259 min Scan# 61
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01

Tgt Ion: 50 Resp: 3958
Ion Ratio Lower Upper
50 100
52 30.8 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#4
Vinyl Chloride
Concen: 21.658 ug/l
RT: 2.412 min Scan# 86
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 62 Resp: 51040
Ion Ratio Lower Upper
62 100
64 30.2 26.8 40.2





#5
Bromomethane
Concen: 20.503 ug/l
RT: 2.826 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion: 94 Resp: 37740
Ion Ratio Lower Upper
94 100
96 95.9 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#6
Chloroethane
Concen: 21.395 ug/l
RT: 2.979 min Scan# 179
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion: 64 Resp: 34249
Ion Ratio Lower Upper
64 100
66 31.0 24.5 36.7





#7
 Trichlorofluoromethane
 Concen: 23.025 ug/l
 RT: 3.308 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51




Instrument :
 MSVOA_W
 ClientSampleId :
 VW0717SBSD01



Tgt Ion:101 Resp: 4927
 Ion Ratio Lower Upper
 101 100
 103 64.1 50.2 75.4




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025



#8
 Diethyl Ether
 Concen: 22.085 ug/l
 RT: 3.722 min Scan# 301
 Delta R.T. 0.000 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51

Tgt Ion: 74 Resp: 35917
 Ion Ratio Lower Upper
 74 100
 45 90.4 43.9 131.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.801 ug/l
RT: 4.100 min Scan# 364
Delta R.T. -0.006 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01

Tgt Ion:101 Resp: 4941
Ion Ratio Lower Upper
101 100
85 44.0 35.6 53.4
151 71.2 56.8 85.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#10
Methyl Iodide
Concen: 20.099 ug/l
RT: 4.308 min Scan# 397
Delta R.T. -0.006 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion:142 Resp: 73606
Ion Ratio Lower Upper
142 100
127 46.8 37.9 56.9
141 14.9 11.7 17.5









#11
Tert butyl alcohol
Concen: 112.964 ug/l
RT: 5.222 min Scan# 546
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01

Tgt Ion: 59 Resp: 2193
Ion Ratio Lower Upper
59 100
57 11.6 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#12
1,1-Dichloroethene
Concen: 21.079 ug/l
RT: 4.076 min Scan# 359
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 96 Resp: 54848
Ion Ratio Lower Upper
96 100
61 161.5 135.8 203.8
98 62.2 50.4 75.6





#13
Acrolein
Concen: 110.454 ug/l
RT: 3.936 min Scan# 336
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion: 56 Resp: 3854
Ion Ratio Lower Upper
56 100
55 74.1 54.0 81.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#14
Allyl chloride
Concen: 21.162 ug/l
RT: 4.710 min Scan# 463
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 41 Resp: 84485
Ion Ratio Lower Upper
41 100
39 71.4 60.7 91.1
76 36.4 30.3 45.5





#15
Acrylonitrile
Concen: 112.726 ug/l
RT: 5.411 min Scan# 51
Delta R.T. 0.013 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion: 53 Resp: 90150
Ion Ratio Lower Upper
53 100
52 83.5 66.4 99.6
51 37.9 29.9 44.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#16
Acetone
Concen: 87.958 ug/l
RT: 4.161 min Scan# 373
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 43 Resp: 68131
Ion Ratio Lower Upper
43 100
58 33.1 24.8 37.2





#17
Carbon Disulfide
Concen: 19.642 ug/l
RT: 4.423 min Scan# 416
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 76 Resp: 136762
Ion Ratio Lower Upper
76 100
78 9.7 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#18
Methyl Acetate
Concen: 16.781 ug/l
RT: 4.716 min Scan# 464
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 43 Resp: 37149
Ion Ratio Lower Upper
43 100
74 25.6 20.6 30.8





#19
Methyl tert-butyl Ether
Concen: 23.034 ug/l
RT: 5.472 min Scan# 518
Delta R.T. 0.013 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 73 Resp: 101720
Ion Ratio Lower Upper
73 100
57 21.3 17.2 25.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#20
Methylene Chloride
Concen: 18.267 ug/l
RT: 4.960 min Scan# 504
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 84 Resp: 65933
Ion Ratio Lower Upper
84 100
49 113.3 98.2 147.4
51 36.8 29.8 44.8
86 62.1 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 21.746 ug/l
RT: 5.460 min Scan# 586
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion: 96 Resp: 6012
Ion Ratio Lower Upper
96 100
61 139.6 109.8 164.6
98 59.8 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#22
Diisopropyl ether
Concen: 23.499 ug/l
RT: 6.344 min Scan# 731
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 45 Resp: 182967
Ion Ratio Lower Upper
45 100
43 61.9 48.3 72.5
87 28.6 22.6 34.0
59 12.7 9.7 14.5





#23
 Vinyl Acetate
 Concen: 123.690 ug/l
 RT: 6.289 min Scan# 715
 Delta R.T. 0.007 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51




Instrument :
 MSVOA_W
 ClientSampleId :
 VW0717SBSD01



Tgt Ion: 43 Resp: 65781
 Ion Ratio Lower Upper
 43 100
 86 11.5 9.2 13.8




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025



#24
 1,1-Dichloroethane
 Concen: 22.045 ug/l
 RT: 6.252 min Scan# 716
 Delta R.T. 0.007 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51


Tgt Ion: 63 Resp: 111283
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.5 10.6
 100 4.9 2.5 7.5





#25
2-Butanone
Concen: 116.343 ug/l
RT: 7.197 min Scan# 81
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion: 43 Resp: 117240
Ion Ratio Lower Upper
43 100
72 28.2 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#26
2,2-Dichloropropane
Concen: 18.088 ug/l
RT: 7.191 min Scan# 870
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 77 Resp: 53272
Ion Ratio Lower Upper
77 100
97 24.8 11.6 34.7





#27
 cis-1,2-Dichloroethene
 Concen: 22.281 ug/l
 RT: 7.191 min Scan# 870
 Delta R.T. 0.000 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51




Instrument :
 MSVOA_W
 ClientSampleId :
 VW0717SBSD01



Tgt Ion: 96 Resp: 7081
 Ion Ratio Lower Upper
 96 100
 61 142.1 0.0 286.8
 98 66.7 0.0 130.4




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025



#28
 Bromochloromethane
 Concen: 21.610 ug/l
 RT: 7.533 min Scan# 926
 Delta R.T. 0.000 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51

Tgt Ion: 49 Resp: 48138
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 2.4
 130 73.9 61.3 91.9





#29

Tetrahydrofuran

Concen: 118.055 ug/l

RT: 7.557 min Scan# 91

Delta R.T. 0.007 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Instrument :

MSVOA_W

ClientSampleId :





VW0717SBS01

Tgt Ion:	42	Resp:	8147
Ion Ratio	Lower	Upper	
42	100		
72	46.5	36.8	55.2
71	43.1	34.6	52.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#30

Chloroform

Concen: 22.357 ug/l




RT: 7.697 min Scan# 953


Delta R.T. 0.007 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51


Tgt Ion:	83	Resp:	119920
Ion Ratio	Lower	Upper	
83	100		
85	60.0	51.4	77.0





#31
Cyclohexane
Concen: 21.935 ug/l
RT: 7.972 min Scan# 998
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 56 Resp: 97870
Ion Ratio Lower Upper
56 100
69 40.9 27.2 40.8
84 92.2 68.5 102.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#32
1,1,1-Trichloroethane
Concen: 20.553 ug/l
RT: 7.892 min Scan# 985
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 97 Resp: 84958
Ion Ratio Lower Upper
97 100
99 63.0 51.0 76.6
61 45.5 36.0 54.0





#33
1,2-Dichloroethane-d4
Concen: 54.388 ug/l
RT: 8.319 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01




Tgt Ion: 65 Resp: 17044
Ion Ratio Lower Upper
65 100
67 51.0 0.0 99.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.856 min Scan# 1143
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion:114 Resp: 418079
Ion Ratio Lower Upper
114 100
63 20.3 0.0 43.6
88 16.8 0.0 34.2





#35
Dibromofluoromethane
Concen: 49.983 ug/l
RT: 7.904 min Scan# 987
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion:113 Resp: 13636
Ion Ratio Lower Upper
113 100
111 105.2 82.1 123.1
192 16.4 13.8 20.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#36
1,1-Dichloropropene
Concen: 20.895 ug/l
RT: 8.100 min Scan# 1019
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion: 75 Resp: 82489
Ion Ratio Lower Upper
75 100
110 33.3 16.8 50.4
77 31.1 25.0 37.6









#37
Ethyl Acetate
Concen: 22.570 ug/l
RT: 7.277 min Scan# 884
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01

Tgt Ion: 43 Resp: 5620
Ion Ratio Lower Upper
43 100
61 13.2 10.9 16.3
70 11.6 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#38
Carbon Tetrachloride
Concen: 19.238 ug/l
RT: 8.087 min Scan# 1017
Delta R.T. 0.006 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion:117 Resp: 77394
Ion Ratio Lower Upper
117 100
119 94.2 75.0 112.6
121 33.5 24.6 37.0





#39
Methylcyclohexane
Concen: 21.443 ug/l
RT: 9.343 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 83 Resp: 10532
Ion Ratio Lower Upper
83 100
55 79.6 61.0 91.4
98 51.4 37.9 56.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#40
Benzene
Concen: 21.404 ug/l
RT: 8.337 min Scan# 1058
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 78 Resp: 249996
Ion Ratio Lower Upper
78 100
77 24.3 19.2 28.8





#41
Methacrylonitrile
Concen: 23.298 ug/l
RT: 7.502 min Scan# 91
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 41 Resp: 3108
Ion Ratio Lower Upper
41 100
39 54.0 43.2 64.8
67 73.0 58.1 87.1
52 29.1 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#42
1,2-Dichloroethane
Concen: 21.397 ug/l
RT: 8.410 min Scan# 1070
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion: 62 Resp: 84445
Ion Ratio Lower Upper
62 100
98 8.9 0.0 17.2





#43
Isopropyl Acetate
Concen: 22.223 ug/l
RT: 8.441 min Scan# 1075
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 43 Resp: 95400
Ion Ratio Lower Upper
43 100
61 30.6 23.1 34.7
87 14.3 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#44
Trichloroethene
Concen: 20.557 ug/l
RT: 9.105 min Scan# 1184
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion:130 Resp: 59927
Ion Ratio Lower Upper
130 100
95 106.7 0.0 208.6





#45
1,2-Dichloropropane
Concen: 21.190 ug/l
RT: 9.374 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01




Tgt Ion: 63 Resp: 59810
Ion Ratio Lower Upper
63 100
65 30.5 24.2 36.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#46
Dibromomethane
Concen: 21.737 ug/l
RT: 9.465 min Scan# 1243
Delta R.T. -0.006 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion: 93 Resp: 39615
Ion Ratio Lower Upper
93 100
95 81.8 67.0 100.4
174 81.5 71.7 107.5





#47
Bromodichloromethane
Concen: 21.016 ug/l
RT: 9.654 min Scan# 1274
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 83 Resp: 89815
Ion Ratio Lower Upper
83 100
85 62.7 50.8 76.2
127 8.2 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#48
Methyl methacrylate
Concen: 21.409 ug/l
RT: 9.447 min Scan# 1240
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion: 41 Resp: 45191
Ion Ratio Lower Upper
41 100
69 91.6 71.6 107.4
39 57.8 49.5 74.3





#49

1,4-Dioxane

Concen: 453.788 ug/l

RT: 9.471 min Scan# 1244

Delta R.T. 0.007 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Instrument :

MSVOA_W

ClientSampleId :

VW0717SBS01

Tgt Ion: 88 Resp: 9680

Ion Ratio Lower Upper

88 100

43 26.5 0.0 0.0





58 74.7 58.3 87.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025

Supervised By :Semsettin Yesilyurt 07/18/2025



#50

Toluene-d8

Concen: 51.655 ug/l

RT: 10.325 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VW031882.D




Acq: 17 Jul 2025 11:51


Tgt Ion: 98 Resp: 524424

Ion Ratio Lower Upper

98 100


100 64.2 53.0 79.4





#51
4-Methyl-2-Pentanone
Concen: 111.209 ug/l
RT: 10.215 min Scan# 1366
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion: 43 Resp: 274460
Ion Ratio Lower Upper
43 100
58 41.3 32.2 48.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#52
Toluene
Concen: 21.466 ug/l
RT: 10.392 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion: 92 Resp: 160528
Ion Ratio Lower Upper
92 100
91 169.2 139.5 209.3





#53
t-1,3-Dichloropropene
Concen: 21.348 ug/l
RT: 10.611 min Scan# 1431
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01




Tgt Ion: 75 Resp: 85600
Ion Ratio Lower Upper
75 100
77 31.2 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#54
cis-1,3-Dichloropropene
Concen: 21.631 ug/l
RT: 10.075 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion: 75 Resp: 98325
Ion Ratio Lower Upper
75 100
77 30.7 25.4 38.2
39 48.2 41.8 62.6





#55
1,1,2-Trichloroethane
Concen: 21.593 ug/l
RT: 10.788 min Scan# 1460
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion: 97	Resp: 51549
Ion Ratio	Lower Upper
97 100	
83 79.1	66.9 100.3
85 54.1	41.1 61.7
99 58.7	46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#56
Ethyl methacrylate
Concen: 23.667 ug/l
RT: 10.648 min Scan# 1437
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion: 69	Resp: 76651
Ion Ratio	Lower Upper
69 100	
41 66.0	53.3 79.9
39 35.7	30.2 45.2





#57
1,3-Dichloropropane
Concen: 21.876 ug/l
RT: 10.934 min Scan# 1484
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion: 76 Resp: 91688
Ion Ratio Lower Upper
76 100
78 32.3 26.2 39.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#58
2-Chloroethyl Vinyl ether
Concen: 102.247 ug/l
RT: 9.928 min Scan# 1319
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion: 63 Resp: 182347
Ion Ratio Lower Upper
63 100
106 29.1 24.1 36.1









#59
2-Hexanone
Concen: 116.188 ug/l
RT: 10.971 min Scan# 1490
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01

Tgt Ion: 43 Resp: 19431
Ion Ratio Lower Upper
43 100
58 55.9 27.9 83.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#60
Dibromochloromethane
Concen: 19.629 ug/l
RT: 11.136 min Scan# 1517
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion:129 Resp: 56037
Ion Ratio Lower Upper
129 100
127 78.9 38.0 114.1





#61
1,2-Dibromoethane
Concen: 21.340 ug/l
RT: 11.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion:107 Resp: 51000
Ion Ratio Lower Upper
107 100
109 97.8 73.9 110.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#62
4-Bromofluorobenzene
Concen: 54.147 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion: 95 Resp: 202301
Ion Ratio Lower Upper
95 100
174 68.7 0.0 133.8
176 60.2 0.0 126.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.636 min Scan# 11
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion:117 Resp: 36785
Ion Ratio Lower Upper
117 100
82 60.6 48.6 72.8
119 30.1 23.9 35.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#64
Tetrachloroethene
Concen: 18.821 ug/l
RT: 10.861 min Scan# 1472
Delta R.T. -0.006 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion:164 Resp: 45287
Ion Ratio Lower Upper
164 100
166 127.3 100.7 151.1
129 115.1 75.4 113.0#
131 107.0 74.0 111.0





#65
Chlorobenzene
Concen: 20.956 ug/l
RT: 11.660 min Scan# 1108
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01




Tgt Ion:112 Resp: 17008
Ion Ratio Lower Upper
112 100
114 33.1 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 21.102 ug/l
RT: 11.727 min Scan# 1614
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion:131 Resp: 54577
Ion Ratio Lower Upper
131 100
133 95.1 48.5 145.6
119 66.0 32.5 97.5





#67
Ethyl Benzene
Concen: 21.531 ug/l
RT: 11.733 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01




Tgt Ion: 91 Resp: 30302
Ion Ratio Lower Upper
91 100
106 30.4 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#68
m/p-Xylenes
Concen: 42.820 ug/l
RT: 11.837 min Scan# 1632
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Tgt Ion:106 Resp: 231665
Ion Ratio Lower Upper
106 100
91 207.8 166.4 249.6





#69
o-Xylene
Concen: 21.856 ug/l
RT: 12.166 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion:106 Resp: 10948
Ion Ratio Lower Upper
106 100
91 221.9 111.5 334.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#70
Styrene
Concen: 22.060 ug/l
RT: 12.178 min Scan# 1688
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion:104 Resp: 191034
Ion Ratio Lower Upper
104 100
78 53.3 42.1 63.1
103 54.6 43.0 64.6





#71

Bromoform

Concen: 19.560 ug/l

RT: 12.349 min Scan# 1716

Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Instrument :

MSVOA_W

ClientSampleId :





VW0717SBS01

Tgt Ion:173	Resp:	30180
Ion Ratio	Lower	Upper
173	100	
175	50.6	24.0 72.0
254	0.0	0.1 0.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.556 min Scan# 1914


Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51


Tgt Ion:152	Resp:	170253
Ion Ratio	Lower	Upper
152	100	
115	69.2	31.9 95.7
150	166.2	0.0 356.4





#73
Isopropylbenzene
Concen: 21.803 ug/l
RT: 12.465 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion:105 Resp: 282352
Ion Ratio Lower Upper
105 100
120 25.7 12.8 38.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#74
N-ethyl acetate
Concen: 25.813 ug/l
RT: 12.270 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51



Tgt Ion: 43 Resp: 90101
Ion Ratio Lower Upper
43 100
70 45.2 36.0 54.0
55 25.4 20.8 31.2
61 25.1 19.7 29.5









#75
1,1,2,2-Tetrachloroethane
Concen: 23.196 ug/l
RT: 12.715 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01

Tgt Ion: 83 Resp: 66690
Ion Ratio Lower Upper
83 100
131 10.6 5.3 16.1
85 63.8 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#76
1,2,3-Trichloropropane
Concen: 23.920 ug/l m
RT: 12.769 min Scan# 1785
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion: 75 Resp: 52628
Ion Ratio Lower Upper
75 100
77 0.0 141.2 423.6#





#77

Bromobenzene

Concen: 21.649 ug/l

RT: 12.745 min Scan# 1781

Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Instrument :

MSVOA_W

ClientSampleId :





VW0717SBS01

Tgt Ion:156 Resp: 62529
 Ion Ratio Lower Upper
 156 100
 77 232.4 117.9 353.7
 158 97.5 52.6 157.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025



#78

n-propylbenzene

Concen: 22.602 ug/l




RT: 12.800 min Scan# 1790


Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Tgt Ion: 91 Resp: 350510
 Ion Ratio Lower Upper
 91 100
 120 21.6 11.5 34.4









#79
2-Chlorotoluene
Concen: 22.851 ug/l
RT: 12.891 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01

Tgt Ion: 91 Resp: 213392
Ion Ratio Lower Upper
91 100
126 31.1 15.8 47.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#80
1,3,5-Trimethylbenzene
Concen: 22.470 ug/l
RT: 12.940 min Scan# 1813
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion:105 Resp: 240871
Ion Ratio Lower Upper
105 100
120 48.5 23.8 71.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.778 ug/l
RT: 12.507 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01



Tgt Ion: 75 Resp: 1962
Ion Ratio Lower Upper
75 100
53 111.0 83.5 125.3
89 48.2 38.3 57.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#82
4-Chlorotoluene
Concen: 22.575 ug/l
RT: 12.983 min Scan# 1820
Delta R.T. -0.006 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion: 91 Resp: 221327
Ion Ratio Lower Upper
91 100
126 31.1 16.0 47.9









#83
tert-Butylbenzene
Concen: 21.391 ug/l
RT: 13.202 min Scan# 1856
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBS01

Tgt Ion:119 Resp: 196418
Ion Ratio Lower Upper
119 100
91 74.9 36.0 108.1
134 25.6 12.4 37.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#84
1,2,4-Trimethylbenzene
Concen: 22.978 ug/l
RT: 13.245 min Scan# 1863
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51


Tgt Ion:105 Resp: 249572
Ion Ratio Lower Upper
105 100
120 42.3 22.1 66.1





#85
 sec-Butylbenzene
 Concen: 22.511 ug/l
 RT: 13.379 min Scan# 1885
 Delta R.T. 0.000 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51




Instrument :
 MSVOA_W
 ClientSampleId :
 VW0717SBSD01




Tgt Ion:105 Resp: 310250
 Ion Ratio Lower Upper
 105 100
 134 18.8 9.6 28.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
 Supervised By :Semsettin Yesilyurt 07/18/2025




#86
 p-Isopropyltoluene
 Concen: 22.121 ug/l
 RT: 13.495 min Scan# 1904
 Delta R.T. 0.000 min
 Lab File: VW031882.D
 Acq: 17 Jul 2025 11:51




Tgt Ion:119 Resp: 251261
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.9 38.7
 91 25.4 12.7 38.0





#87
1,3-Dichlorobenzene
Concen: 22.206 ug/l
RT: 13.495 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01




Tgt Ion:146 Resp: 12862
Ion Ratio Lower Upper
146 100
111 44.2 22.3 66.9
148 64.7 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025




#88
1,4-Dichlorobenzene
Concen: 21.475 ug/l
RT: 13.574 min Scan# 1917
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51



Tgt Ion:146 Resp: 127323
Ion Ratio Lower Upper
146 100
111 44.0 20.4 61.3
148 63.7 30.2 90.6









#89
n-Butylbenzene
Concen: 22.326 ug/l
RT: 13.818 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01

Tgt Ion: 91 Resp: 246392
Ion Ratio Lower Upper
91 100
92 54.0 27.2 81.5
134 23.6 12.2 36.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#90
Hexachloroethane
Concen: 20.693 ug/l
RT: 14.086 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion:117 Resp: 42404
Ion Ratio Lower Upper
117 100
201 58.5 31.1 93.3





#91

1,2-Dichlorobenzene

Concen: 22.129 ug/l

RT: 13.867 min Scan# 1965

Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Instrument :

MSVOA_W

ClientSampleId :

VW0717SBS01

Tgt Ion:146 Resp: 11715

Ion Ratio Lower Upper

146 100

111 47.0 21.7 65.1





148 65.0 29.8 89.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025

Supervised By :Semsettin Yesilyurt 07/18/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 22.085 ug/l

RT: 14.483 min Scan# 2066

Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51




Tgt Ion: 75 Resp: 12398


Ion Ratio Lower Upper

75 100

155 67.9 37.2 111.6

157 91.6 45.6 136.9





#93

1,2,4-Trichlorobenzene

Concen: 20.943 ug/l

RT: 15.129 min Scan# 2172

Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51

Instrument :

MSVOA_W

ClientSampleId :

VW0717SBS01

Tgt Ion:180 Resp: 6838

Ion Ratio Lower Upper

180 100

182 97.7 48.6 145.9





145 36.0 17.3 51.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025

Supervised By :Semsettin Yesilyurt 07/18/2025



#94

Hexachlorobutadiene

Concen: 19.216 ug/l

RT: 15.226 min Scan# 2188

Delta R.T. 0.000 min

Lab File: VW031882.D

Acq: 17 Jul 2025 11:51




Tgt Ion:225 Resp: 30297


Ion Ratio Lower Upper

225 100

223 62.0 32.6 97.8


227 61.4 37.3 111.8





#95
Naphthalene
Concen: 23.612 ug/l
RT: 15.360 min Scan# 2210
Delta R.T. 0.007 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51




Instrument :
MSVOA_W
ClientSampleId :
VW0717SBSD01



Tgt Ion:128 Resp: 189000
Ion Ratio Lower Upper
128 100
127 13.1 10.8 16.2
129 11.0 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/18/2025
Supervised By :Semsettin Yesilyurt 07/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.901 ug/l
RT: 15.549 min Scan# 2241
Delta R.T. 0.000 min
Lab File: VW031882.D
Acq: 17 Jul 2025 11:51

Tgt Ion:180 Resp: 59707
Ion Ratio Lower Upper
180 100
182 97.0 46.4 139.1
145 41.4 17.8 53.3

