

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW071919\
Quantitation Report (QT Reviewed)


Data File : VW011380.D
Acq On : 19 Jul 2019 16:15
Operator : SY/VA
Sample : K3863-14
Misc : 6.16G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A52F4

Manual Integrations
APPROVED

MMDadoda
7/24/2019 12:49:01 AM

Quant Time: Jul 20 01:56:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 20 00:46:56 2019
Response via : Initial Calibration



Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW071919\
Quantitation Report (Qedit)




Data File : VW011380.D
Acq On : 19 Jul 2019 16:15
Operator : SY/VA
Sample : K3863-14
Misc : 6.16G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A52F4

Manual Integrations
APPROVED

MMDadoda
7/24/2019 12:49:01 AM

Quant Time: Jul 20 00:51:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 20 00:46:56 2019
Response via : Initial Calibration



TIC: VW011380.D

(15) Methyl Acetate (T)

4.660min (+0.000) 0.54ug/L

response 1673

Ion	Exp%	Act%
43.00	100	100
74.00	18.90	43.22#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW071919\
Quantitation Report (Qedit)




Data File : VW011380.D
Acq On : 19 Jul 2019 16:15
Operator : SY/VA
Sample : K3863-14
Misc : 6.16G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A52F4

Manual Integrations
APPROVED

MMDadoda
7/24/2019 12:49:01 AM

Quant Time: Jul 20 00:51:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 20 00:46:56 2019
Response via : Initial Calibration



TIC: VW011380.D

(15) Methyl Acetate (T)

4.660min (+0.000) 1.52ug/L m

response 4712

Ion	Exp%	Act%
43.00	100	100
74.00	18.90	15.34
0.00	0.00	0.00
0.00	0.00	0.00

507/22/19 Sy

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW071919\
Quantitation Report (QT Reviewed)

Data File : VW011380.D
Acq On : 19 Jul 2019 16:15
Operator : SY/VA
Sample : K3863-14
Misc : 6.16G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A52F4

Manual Integrations
APPROVED

MMDadoda
7/24/2019 12:49:01 AM

Quant Time: Jul 20 01:56:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 20 00:46:56 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	323879	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	269999	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	102113	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	56057	11.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	45.32%
7) Chloroethane-d5	2.89	69	50175	14.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.84%
10) 1,1-Dichloroethene-d2	4.02	63	113286	10.82	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	43.28%#
20) 2-Butanone-d5	7.07	46	85427	43.85	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.70%
24) Chloroform-d	7.65	84	156872	17.67	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	70.68%
26) 1,2-Dichloroethane-d4	8.31	65	101285	18.53	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	74.12%
29) Benzene-d6	8.27	84	304564	18.31	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.24%
33) 1,2-Dichloropropane-d6	9.27	67	108687	20.07	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.28%
37) Toluene-d8	10.32	98	245355	16.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	66.20%
38) trans-1,3-Dichloropropene-	10.58	79	40050	17.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.56%
39) 2-Hexanone-d5	10.93	63	58427	47.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.08%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	91619	20.45	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.80%
61) 1,2-Dichlorobenzene-d4	13.85	152	70350	17.95	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	71.80%#

Target Compounds

					Qvalue	
12) 1,1-Dichloroethene	4.04	96	11395	2.677	ug/L	# 1
13) Acetone	4.12	43	61637	32.392	ug/L	98
14) Carbon disulfide	4.38	76	26309	2.161	ug/L	95
15) Methyl Acetate	4.66	43	4712m)	1.522	ug/L	94
19) 1,1-Dichloroethane	6.21	63	16612	1.903	ug/L	94
21) 2-Butanone	7.18	43	7149	2.910	ug/L	78
22) cis-1,2-Dichloroethene	7.17	96	12065	2.592	ug/L	99
31) 1,1,1-Trichloroethane	7.87	97	208105	37.137	ug/L	100
35) Trichloroethene	9.09	95	12485	2.932	ug/L	97
47) Tetrachloroethene	10.86	164	16229	5.043	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed