

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072122\
 Data File : VW023967.D
 Acq On : 21 Jul 2022 08:58
 Operator : SY/VA
 Sample : VW0721WBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jul 22 01:29:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 22 01:28:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	266083	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	256513	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	113489	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	128018	18.837	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.360%
7) Chloroethane-d5	2.873	69	82538	18.205	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.840%
11) 1,1-Dichloroethene-d2	4.196	63	199	0.036	ug/L	0.19
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.160%#
21) 2-Butanone-d5	7.074	46	55022	52.197	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.400%
24) Chloroform-d	7.647	84	188253	24.300	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	8.305	65	106659	22.511	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.040%
32) Benzene-d6	8.268	84	342341	25.674	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.680%
36) 1,2-Dichloropropane-d6	9.274	67	104521	26.022	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.080%
41) Toluene-d8	10.323	98	312734	23.122	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.480%
43) trans-1,3-Dichloroprop...	10.573	79	43394	23.049	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.200%
47) 2-Hexanone-d5	10.920	63	44061	52.168	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.340%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	105383	23.741	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.960%
66) 1,2-Dichlorobenzene-d4	13.847	152	114455	28.251	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	113.000%
Target Compounds					Qvalue	
48) 2-Hexanone	10.920	43	5619	3.035	ug/L #	70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072122\
Data File : VW023967.D
Acq On : 21 Jul 2022 08:58
Operator : SY/VA
Sample : VW0721WBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :


Quant Time: Jul 22 01:29:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 22 01:28:21 2022
Response via : Initial Calibration





#48
 2-Hexanone
 Concen: 3.035 ug/L
 RT: 10.920 min Scan# 1486
 Delta R.T. -0.043 min
 Lab File: VW023967.D
 Acq: 21 Jul 2022 08:58

Instrument :
 MSVOA_W
 ClientSampleId :



Tgt Ion: 43	Resp:	5619	
Ion	Ratio	Lower	Upper
43	100		
58	34.1	41.6	62.4#
57	39.2	14.6	21.8#
100	4.2	10.6	15.8#

