

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072222\
 Data File : VW024036.D
 Acq On : 23 Jul 2022 06:57
 Operator : SY/VA
 Sample : N3795-01
 Misc : 6.50g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jul 24 00:40:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jul 23 06:00:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	260183	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	246731	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	101085	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	105743	15.912	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.640%
7) Chloroethane-d5	2.879	69	67094	15.134	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.520%
11) 1,1-Dichloroethene-d2	4.001	63	67445	12.464	ug/L	-0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	49.840%
21) 2-Butanone-d5	7.074	46	32646	31.672	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.340%
24) Chloroform-d	7.647	84	153594	20.276	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.120%
26) 1,2-Dichloroethane-d4	8.305	65	83606	18.046	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	72.200%
32) Benzene-d6	8.269	84	267785	20.879	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.520%
36) 1,2-Dichloropropane-d6	9.268	67	86126	22.293	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.160%
41) Toluene-d8	10.323	98	246166	18.922	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.680%
43) trans-1,3-Dichloroprop...	10.573	79	30626	16.912	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.640%
47) 2-Hexanone-d5	10.927	63	25039	30.821	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.640%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	72805	17.052	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	68.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	79099	21.920	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072222\
Data File : VW024036.D
Acq On : 23 Jul 2022 06:57
Operator : SY/VA
Sample : N3795-01
Misc : 6.50g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jul 24 00:40:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jul 23 06:00:17 2022
Response via : Initial Calibration

