

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072222\
 Data File : VW024047.D
 Acq On : 23 Jul 2022 11:30
 Operator : SY/VA
 Sample : N3738-08
 Misc : 5.01g/10mL/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jul 24 00:41:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jul 23 06:00:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	207493	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	218164	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	99686	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	102209	19.286	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.160%
7) Chloroethane-d5	2.873	69	64726	18.308	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.240%
11) 1,1-Dichloroethene-d2	4.013	63	58099	13.463	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.840%
21) 2-Butanone-d5	7.080	46	40158	48.853	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.700%
24) Chloroform-d	7.647	84	126079	20.870	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.480%
26) 1,2-Dichloroethane-d4	8.470	65	98	0.027	ug/L	0.16
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.120%#
32) Benzene-d6	8.275	84	223234	19.684	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.720%
36) 1,2-Dichloropropane-d6	9.274	67	74343	21.763	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.040%
41) Toluene-d8	10.323	98	206321	17.936	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	71.760%
43) trans-1,3-Dichloroprop...	10.573	79	29755	18.582	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.320%
47) 2-Hexanone-d5	10.920	63	32351	45.036	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.080%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	79242	20.990	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.960%
66) 1,2-Dichlorobenzene-d4	13.853	152	79889	22.450	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072222\
Data File : VW024047.D
Acq On : 23 Jul 2022 11:30
Operator : SY/VA
Sample : N3738-08
Misc : 5.01g/10mL/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jul 24 00:41:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jul 23 06:00:17 2022
Response via : Initial Calibration

