

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072222\
 Data File : VW024015.D
 Acq On : 22 Jul 2022 19:17
 Operator : SY/VA
 Sample : VW0722SBL02
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK560

Quant Time: Jul 23 03:31:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jul 23 03:31:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	207087	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	205749	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	91208	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	120418	22.767	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.080%
7) Chloroethane-d5	2.879	69	77702	22.021	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.080%
11) 1,1-Dichloroethene-d2	4.007	63	71501	16.601	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.400%
21) 2-Butanone-d5	7.080	46	53035	64.645	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.280%
24) Chloroform-d	7.647	84	150765	25.006	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.040%
26) 1,2-Dichloroethane-d4	8.305	65	91110	24.707	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.840%
32) Benzene-d6	8.275	84	270837	25.323	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.280%
36) 1,2-Dichloropropane-d6	9.274	67	86861	26.961	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	107.840%
41) Toluene-d8	10.323	98	242503	22.353	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	10.573	79	34506	22.850	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.400%
47) 2-Hexanone-d5	10.927	63	41158	60.754	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.500%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	92983	26.116	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.480%
66) 1,2-Dichlorobenzene-d4	13.847	152	86991	26.718	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072222\
Data File : VW024015.D
Acq On : 22 Jul 2022 19:17
Operator : SY/VA
Sample : VW07222SBL02
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK560

Quant Time: Jul 23 03:31:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jul 23 03:31:10 2022
Response via : Initial Calibration

