

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072224\
 Data File : VW029686.D
 Acq On : 22 Jul 2024 16:14
 Operator : SY/MD
 Sample : P3278-02
 Misc : 7.29g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E1ZP7

Quant Time: Jul 23 04:10:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 19 23:16:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	321212	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	268107	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	95650	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	46998	9.057	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	36.240%
7) Chloroethane-d5	2.893	69	47036	12.589	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.360%
11) 1,1-Dichloroethene-d2	4.021	65	21653	9.293	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	37.160%#
21) 2-Butanone-d5	7.081	46	44682	28.685	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	57.380%
24) Chloroform-d	7.648	84	172082	18.531	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	74.120%
26) 1,2-Dichloroethane-d4	8.307	65	107200	19.682	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.720%
32) Benzene-d6	8.276	84	260062	16.192	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.760%
36) 1,2-Dichloropropane-d6	9.276	67	105955	20.633	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.520%
41) Toluene-d8	10.319	98	204989	14.513	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	58.040%
43) trans-1,3-Dichloroprop...	10.575	79	35356	16.259	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.040%
47) 2-Hexanone-d5	10.922	63	31714	33.986	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	67.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	94775	22.731	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.920%
66) 1,2-Dichlorobenzene-d4	13.848	152	70094	21.042	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.160%
Target Compounds						
16) Methylene chloride	4.923	84	21218	3.690	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072224\
Data File : VW029686.D
Acq On : 22 Jul 2024 16:14
Operator : SY/MD
Sample : P3278-02
Misc : 7.29g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E1ZP7

Quant Time: Jul 23 04:10:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 19 23:16:32 2024
Response via : Initial Calibration

