

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072224\
 Data File : VW029690.D
 Acq On : 22 Jul 2024 17:50
 Operator : SY/MD
 Sample : P3278-12
 Misc : 10.84g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E1ZR0

Quant Time: Jul 23 04:11:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 19 23:16:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	237660	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	203511	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	76819	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	51056	13.298	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.200%
7) Chloroethane-d5	2.899	69	40961	14.817	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.280%
11) 1,1-Dichloroethene-d2	4.021	65	24838	14.407	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.640%
21) 2-Butanone-d5	7.081	46	39858	34.584	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.160%
24) Chloroform-d	7.648	84	116972	17.025	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	68.120%
26) 1,2-Dichloroethane-d4	8.307	65	73303	18.190	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	72.760%
32) Benzene-d6	8.276	84	209461	17.181	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.720%
36) 1,2-Dichloropropane-d6	9.270	67	73228	18.786	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	75.160%
41) Toluene-d8	10.319	98	174433	16.269	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	65.080%
43) trans-1,3-Dichloroprop...	10.575	79	25945	15.718	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	62.880%
47) 2-Hexanone-d5	10.922	63	24427	34.486	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	62477	19.741	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.960%
66) 1,2-Dichlorobenzene-d4	13.848	152	48181	18.010	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.040%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072224\
Data File : VW029690.D
Acq On : 22 Jul 2024 17:50
Operator : SY/MD
Sample : P3278-12
Misc : 10.84g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E1ZR0

Quant Time: Jul 23 04:11:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 19 23:16:32 2024
Response via : Initial Calibration

