

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072324\
 Data File : VW029703.D
 Acq On : 23 Jul 2024 14:18
 Operator : SY/MD
 Sample : P3278-18
 Misc : 9.57g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E1ZS0

Quant Time: Jul 24 01:15:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 24 01:12:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	302743	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	268452	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	92931	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	56865	11.627	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	46.520%		
7) Chloroethane-d5	2.893	69	49107	13.945	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	55.760%		
11) 1,1-Dichloroethene-d2	4.021	65	27657	12.593	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	50.360%		
21) 2-Butanone-d5	7.075	46	49585	33.775	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	67.560%		
24) Chloroform-d	7.648	84	68477	7.824	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	31.280%#		
26) 1,2-Dichloroethane-d4	8.307	65	91137	17.753	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	71.000%		
32) Benzene-d6	8.276	84	258752	16.090	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	64.360%		
36) 1,2-Dichloropropane-d6	9.270	67	93757	18.234	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	72.920%		
41) Toluene-d8	10.325	98	209717	14.828	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	59.320%		
43) trans-1,3-Dichloroprop...	10.575	79	32315	14.841	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	59.360%		
47) 2-Hexanone-d5	10.922	63	31276	33.474	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	66.940%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	9837	2.356	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	9.440%#		
66) 1,2-Dichlorobenzene-d4	13.849	152	63367	19.580	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	78.320%		
Target Compounds					Qvalue	
13) Acetone	4.119	43	9277	6.261	ug/L	96
16) Methylene chloride	4.917	84	13144	2.425	ug/L	97
25) Chloroform	7.679	83	80289	9.695	ug/L	95
34) Trichloroethene	9.093	95	3515	0.792	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072324\
Data File : VW029703.D
Acq On : 23 Jul 2024 14:18
Operator : SY/MD
Sample : P3278-18
Misc : 9.57g/10mL/MSVOA_W/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E1ZS0

Quant Time: Jul 24 01:15:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 24 01:12:32 2024
Response via : Initial Calibration





#13

Acetone

Concen: 6.261 ug/L

RT: 4.119 min Scan# 3

Delta R.T. -0.012 min

Lab File: VW029703.D

Acq: 23 Jul 2024 14:18

Instrument :

MSVOA_W

ClientSampleId :





E1ZS0

Tgt Ion: 43 Resp: 9277

Ion Ratio Lower Upper

43 100

58 29.7 0.0 55.6



#16

Methylene chloride

Concen: 2.425 ug/L

RT: 4.917 min Scan# 497

Delta R.T. 0.000 min

Lab File: VW029703.D

Acq: 23 Jul 2024 14:18




Tgt Ion: 84 Resp: 13144


Ion Ratio Lower Upper

84 100

86 63.0 46.2 85.8

49 141.7 96.7 179.7









#25
Chloroform
Concen: 9.695 ug/L
RT: 7.679 min Scan# 91
Delta R.T. 0.000 min
Lab File: VW029703.D
Acq: 23 Jul 2024 14:18

Instrument :
MSVOA_W
ClientSampleId :
E1ZS0

Tgt Ion: 83 Resp: 80289
Ion Ratio Lower Upper
83 100
85 66.9 44.1 81.9



#34
Trichloroethene
Concen: 0.792 ug/L
RT: 9.093 min Scan# 1182
Delta R.T. 0.000 min
Lab File: VW029703.D
Acq: 23 Jul 2024 14:18

Tgt Ion: 95 Resp: 3515
Ion Ratio Lower Upper
95 100
97 69.5 47.3 87.9
132 99.0 65.8 122.2
130 94.6 67.8 125.8

