

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072325\
 Data File : VW031904.D
 Acq On : 23 Jul 2025 11:52
 Operator : SY/MD
 Sample : Q2668-03
 Misc : 4.23g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-2-VOC

Quant Time: Jul 24 02:50:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 23 08:18:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	184054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	393072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	342365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	145319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	142208	51.076	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	102.160%	
35) Dibromofluoromethane	7.898	113	124696	46.686	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	93.380%	
50) Toluene-d8	10.324	98	456182	45.127	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	90.260%	
62) 4-Bromofluorobenzene	12.617	95	155846	40.150	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	80.300%	
Target Compounds						
16) Acetone	4.155	43	12460	18.859	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072325\
Data File : VW031904.D
Acq On : 23 Jul 2025 11:52
Operator : SY/MD
Sample : Q2668-03
Misc : 4.23g/5mL/MSVOA_W/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-2-VOC

Quant Time: Jul 24 02:50:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 23 08:18:46 2025
Response via : Initial Calibration

