

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072418\
 Data File : VW004171.D
 Acq On : 24 Jul 2018 13:17
 Operator : SY/AP
 Sample : J4040-02
 Misc : 5.23G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AD9

Quant Time: Jul 24 17:59:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM071818S.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jul 23 15:53:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	311124	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	312525	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	144326	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	59677	19.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.56%
7) Chloroethane-d5	2.89	69	47273	21.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.24%
10) 1,1-Dichloroethene-d2	4.00	63	113150	15.33	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.32%
20) 2-Butanone-d5	7.09	46	87146	48.29	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.58%
24) Chloroform-d	7.65	84	192769	26.10	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.40%
26) 1,2-Dichloroethane-d4	8.31	65	116530	26.20	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.80%
29) Benzene-d6	8.28	84	335982	23.18	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.72%
33) 1,2-Dichloropropane-d6	9.28	67	124192	25.68	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.72%
37) Toluene-d8	10.33	98	291065	22.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.36%
38) trans-1,3-Dichloropropene-	10.58	79	49675	23.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.52%
39) 2-Hexanone-d5	10.93	63	60274	46.31	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.62%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	131382	24.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.20%
61) 1,2-Dichlorobenzene-d4	13.86	152	127315	27.43	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	109.72%

Target Compounds					Ovalue
47) Tetrachloroethene	10.87	164	41498	10.04 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072418\
Data File : VW004171.D
Acq On : 24 Jul 2018 13:17
Operator : SY/AP
Sample : J4040-02
Misc : 5.23G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AD9

Quant Time: Jul 24 17:59:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM071818S.M
Quant Title : VOC Analysis
QLast Update : Mon Jul 23 15:53:12 2018
Response via : Initial Calibration

