

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072423\
 Data File : VW026642.D
 Acq On : 24 Jul 2023 11:47
 Operator : SY/MD
 Sample : 03712-10
 Misc : 3.94g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH2R8

Quant Time: Jul 25 01:54:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 25 01:52:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	828712	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	751240	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	353233	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	167094	11.953	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	47.800%	
7) Chloroethane-d5	2.893	69	153930	15.693	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	62.760%	
11) 1,1-Dichloroethene-d2	4.021	65	81119	12.357	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	49.440%	
21) 2-Butanone-d5	7.075	46	178197	39.444	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	78.880%	
24) Chloroform-d	7.648	84	513664	21.529	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	86.120%	
26) 1,2-Dichloroethane-d4	8.306	65	304417	22.851	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	91.400%	
32) Benzene-d6	8.270	84	929974	18.033	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	72.120%	
36) 1,2-Dichloropropane-d6	9.270	67	345001	21.757	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	87.040%	
41) Toluene-d8	10.324	98	746485	16.877	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	67.520%	
43) trans-1,3-Dichloroprop...	10.580	79	126125	19.848	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	79.400%	
47) 2-Hexanone-d5	10.922	63	108825	35.171	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	70.340%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	271375	20.209	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	80.840%	
66) 1,2-Dichlorobenzene-d4	13.854	152	292862	22.582	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	90.320%	
Target Compounds						
					Qvalue	
13) Acetone	4.124	43	363239	111.777	ug/L	96
16) Methylene chloride	4.917	84	42892	2.930	ug/L	94
25) Chloroform	7.672	83	16923	0.778	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072423\
Data File : VW026642.D
Acq On : 24 Jul 2023 11:47
Operator : SY/MD
Sample : 03712-10
Misc : 3.94g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH2R8

Quant Time: Jul 25 01:54:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071123SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 25 01:52:41 2023
Response via : Initial Calibration

