

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072425\
 Data File : VW031923.D
 Acq On : 24 Jul 2025 12:07
 Operator : SY/MD
 Sample : Q2689-01
 Misc : 5.02g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-03-07232025

Quant Time: Jul 25 03:43:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 23 08:18:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.971	168	182835	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	384822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	362039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	174319	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	145756	52.699	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	105.400%
35) Dibromofluoromethane	7.898	113	125999	48.185	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	96.360%
50) Toluene-d8	10.331	98	479484	48.449	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.617	95	171852	45.223	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	90.440%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072425\
Data File : VW031923.D
Acq On : 24 Jul 2025 12:07
Operator : SY/MD
Sample : Q2689-01
Misc : 5.02g/5mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-03-07232025

Quant Time: Jul 25 03:43:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 23 08:18:46 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.971 min Scan# 998
Delta R.T. 0.006 min
Lab File: VW031923.D
Acq: 24 Jul 2025 12:07




Instrument :
MSVOA_W
ClientSampleId :
OR-03-07232025


Tgt Ion:168 Resp: 182835
Ion Ratio Lower Upper
168 100
99 64.0 48.2 72.2



#33
1,2-Dichloroethane-d4
Concen: 52.699 ug/l
RT: 8.319 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VW031923.D
Acq: 24 Jul 2025 12:07

Tgt Ion: 65 Resp: 145756
Ion Ratio Lower Upper
65 100
67 51.1 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.849 min Scan# 1143

Delta R.T. -0.006 min

Lab File: VW031923.D

Acq: 24 Jul 2025 12:07

Instrument :

MSVOA_W

ClientSampleId :

OR-03-07232025





Tgt Ion:114 Resp: 384822

Ion Ratio Lower Upper

114 100

63 21.0 0.0 44.2

88 16.3 0.0 35.6



#35

Dibromofluoromethane

Concen: 48.185 ug/l

RT: 7.898 min Scan# 986

Delta R.T. -0.006 min

Lab File: VW031923.D

Acq: 24 Jul 2025 12:07




Tgt Ion:113 Resp: 125999


Ion Ratio Lower Upper

113 100

111 101.9 79.9 119.9

192 16.0 12.6 19.0









#50
Toluene-d8
Concen: 48.449 ug/l
RT: 10.331 min Scan# 11
Delta R.T. 0.006 min
Lab File: VW031923.D
Acq: 24 Jul 2025 12:07




Instrument :
MSVOA_W
ClientSampleId :
OR-03-07232025


Tgt Ion: 98 Resp: 479484
Ion Ratio Lower Upper
98 100
100 64.6 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 45.223 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. 0.000 min
Lab File: VW031923.D
Acq: 24 Jul 2025 12:07

Tgt Ion: 95 Resp: 171852
Ion Ratio Lower Upper
95 100
174 63.3 0.0 129.8
176 62.3 0.0 130.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.629 min Scan# 11
Delta R.T. -0.006 min
Lab File: VW031923.D
Acq: 24 Jul 2025 12:07

Instrument :
MSVOA_W
ClientSampleId :
OR-03-07232025

Tgt Ion:117 Resp: 362039
Ion Ratio Lower Upper
117 100
82 58.1 46.5 69.7
119 30.1 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.556 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VW031923.D
Acq: 24 Jul 2025 12:07

Tgt Ion:152 Resp: 174319
Ion Ratio Lower Upper
152 100
115 65.7 33.4 100.1
150 159.7 0.0 352.2

