

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072622\
 Data File : VW024091.D
 Acq On : 26 Jul 2022 10:55
 Operator : SY/VA
 Sample : N3766-05
 Misc : 5.08g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jul 26 23:40:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 26 23:39:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	225450	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	236623	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	103191	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	124680	25.800	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.200%
7) Chloroethane-d5	2.971	69	252	0.077	ug/L	0.10
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.320%#
11) 1,1-Dichloroethene-d2	4.007	63	72606	17.063	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.240%
21) 2-Butanone-d5	7.074	46	39283	45.739	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.480%
24) Chloroform-d	7.647	84	134594	21.715	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.880%
26) 1,2-Dichloroethane-d4	8.305	65	81100	22.775	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.080%
32) Benzene-d6	8.269	84	264024	24.020	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.080%
36) 1,2-Dichloropropane-d6	9.274	67	76273	22.000	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.000%
41) Toluene-d8	10.323	98	243821	22.180	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.720%
43) trans-1,3-Dichloroprop...	10.579	79	32268	20.991	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.960%
47) 2-Hexanone-d5	10.920	63	31030	42.811	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.620%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	76268	19.065	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	76.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	83420	24.962	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072622\
Data File : VW024091.D
Acq On : 26 Jul 2022 10:55
Operator : SY/VA
Sample : N3766-05
Misc : 5.08g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jul 26 23:40:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 26 23:39:35 2022
Response via : Initial Calibration

