

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072622\
 Data File : VW024093.D
 Acq On : 26 Jul 2022 11:43
 Operator : SY/VA
 Sample : N3741-13
 Misc : 5.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jul 26 23:41:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 26 23:39:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	206951	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	212008	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	88455	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	124088	27.973	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	111.880%
7) Chloroethane-d5	2.873	69	80198	26.757	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.040%
11) 1,1-Dichloroethene-d2	4.007	63	70042	17.932	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.720%
21) 2-Butanone-d5	7.080	46	35122	44.549	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.100%
24) Chloroform-d	7.647	84	134713	23.678	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.720%
26) 1,2-Dichloroethane-d4	8.305	65	78841	24.120	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.480%
32) Benzene-d6	8.275	84	255812	25.975	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.880%
36) 1,2-Dichloropropane-d6	9.275	67	76236	24.542	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.160%
41) Toluene-d8	10.323	98	236045	23.966	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.880%
43) trans-1,3-Dichloroprop...	10.579	79	30496	22.142	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.560%
47) 2-Hexanone-d5	10.927	63	27712	42.672	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.340%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	71906	20.062	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.240%
66) 1,2-Dichlorobenzene-d4	13.847	152	82911	28.942	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	115.760%
Target Compounds						Qvalue
16) Methylene chloride	4.916	84	18961	6.020	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072622\
Data File : VW024093.D
Acq On : 26 Jul 2022 11:43
Operator : SY/VA
Sample : N3741-13
Misc : 5.03g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jul 26 23:41:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 26 23:39:35 2022
Response via : Initial Calibration

