

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW073125\
 Data File : VW031968.D
 Acq On : 31 Jul 2025 14:17
 Operator : SY/MD
 Sample : Q2734-01
 Misc : 5.00g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VNJ-253

Quant Time: Aug 01 01:15:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 23 08:18:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	156479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	340992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	329022	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	147315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	138808	58.640	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	117.280%	
35) Dibromofluoromethane	7.898	113	116885	50.445	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	100.880%	
50) Toluene-d8	10.325	98	428687	48.884	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	97.760%	
62) 4-Bromofluorobenzene	12.617	95	148884	44.215	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	88.420%	
Target Compounds						
16) Acetone	4.155	43	7613	13.553	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW073125\
Data File : VW031968.D
Acq On : 31 Jul 2025 14:17
Operator : SY/MD
Sample : Q2734-01
Misc : 5.00g/5mL/MSVOA_W/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VNJ-253

Quant Time: Aug 01 01:15:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 23 08:18:46 2025
Response via : Initial Calibration

