

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080125\
 Data File : VW031977.D
 Acq On : 01 Aug 2025 11:56
 Operator : SY/MD
 Sample : Q2743-01
 Misc : 5.16g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WC1

Quant Time: Aug 01 23:11:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 23 08:18:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	174433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	365996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.636	117	361399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.550	152	177152	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	143395	54.342	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	108.680%
35) Dibromofluoromethane	7.904	113	125146	50.320	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	100.640%
50) Toluene-d8	10.325	98	454906	48.330	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	96.660%
62) 4-Bromofluorobenzene	12.617	95	164274	45.452	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	90.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080125\
Data File : VW031977.D
Acq On : 01 Aug 2025 11:56
Operator : SY/MD
Sample : Q2743-01
Misc : 5.16g/5mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WC1

Quant Time: Aug 01 23:11:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 23 08:18:46 2025
Response via : Initial Calibration

