

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080218\
 Data File : VW004408.D
 Acq On : 02 Aug 2018 19:02
 Operator : SY/AP
 Sample : J4209-11
 Misc : 7.60G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AC7

Quant Time: Aug 03 04:49:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 02 17:07:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	841904	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	790507	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	281168	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	336314	32.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	128.48%
7) Chloroethane-d5	2.89	69	268134	30.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	121.76%
10) 1,1-Dichloroethene-d2	4.01	63	468875	22.05	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.20%
20) 2-Butanone-d5	7.09	46	173477	37.57	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.14%
24) Chloroform-d	7.65	84	615998	27.03	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	108.12%
26) 1,2-Dichloroethane-d4	8.31	65	360952	28.41	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	113.64%
29) Benzene-d6	8.28	84	1209026	28.75	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	115.00%
33) 1,2-Dichloropropane-d6	9.28	67	395910	28.81	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	115.24%
37) Toluene-d8	10.33	98	994483	26.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.28%
38) trans-1,3-Dichloropropene-	10.58	79	142509	25.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	100.84%
39) 2-Hexanone-d5	10.93	63	132739	39.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.34%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	332650	22.32	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.28%
61) 1,2-Dichlorobenzene-d4	13.86	152	286667	27.15	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.60%


Target Compounds					Ovalue
59) 1,2,3-Trichloropropane	13.57	75	35773	3.288 ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW080218\
Data File : VW004408.D
Acq On : 02 Aug 2018 19:02
Operator : SY/AP
Sample : J4209-11
Misc : 7.60G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AC7

Quant Time: Aug 03 04:49:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 02 17:07:44 2018
Response via : Initial Calibration





#59

1,2,3-Trichloropropane

Concen: 3.288 ug/L

RT: 13.57 min Scan# 1965

Delta R.T. 0.79 min

Lab File: VW004408.D

Acq: 02 Aug 2018 19:02

Instrument :

MSVOA_W

ClientSampleId :

C0AC7

Tot Ion: 75 Resp: 35773

Ion Ratio Lower Upper

75 100

77 34.9 25.3 37.9

