

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW080218\
 Data File : VW004403.D
 Acq On : 02 Aug 2018 15:25
 Operator : SY/AP
 Sample : VIBLK91
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK91

Quant Time: Aug 03 04:49:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 02 17:07:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	1040011	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	1037095	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	468457	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	255819	19.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.12%
7) Chloroethane-d5	2.89	69	217213	19.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.84%
10) 1,1-Dichloroethene-d2	4.01	63	383680	14.61	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.44%
20) 2-Butanone-d5	7.09	46	235894	41.36	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.72%
24) Chloroform-d	7.65	84	564155	20.04	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.16%
26) 1,2-Dichloroethane-d4	8.31	65	332313	21.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.68%
29) Benzene-d6	8.28	84	1107815	20.08	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.32%
33) 1,2-Dichloropropane-d6	9.27	67	366184	20.31	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.24%
37) Toluene-d8	10.33	98	967554	19.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.56%
38) trans-1,3-Dichloropropene-	10.58	79	146390	19.74	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.96%
39) 2-Hexanone-d5	10.93	63	178742	40.72	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.44%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	387675	19.83	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.32%
61) 1,2-Dichlorobenzene-d4	13.86	152	396099	22.51	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.04%

Target Compounds					Qvalue
13) Acetone	4.14	43	18287	3.701 ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW080218\
Data File : VW004403.D
Acq On : 02 Aug 2018 15:25
Operator : SY/AP
Sample : VIBLK91
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK91

Quant Time: Aug 03 04:49:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 02 17:07:44 2018
Response via : Initial Calibration

