

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW080218\
 Data File : VW004446.D
 Acq On : 03 Aug 2018 14:13
 Operator : SY/AP
 Sample : J4334-01
 Misc : 5.45G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BEY64

Quant Time: Aug 04 02:12:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 04 01:26:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	900962	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	793693	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	221811	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	310485	27.71	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.84%
7) Chloroethane-d5	2.89	69	239583	25.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.64%
10) 1,1-Dichloroethene-d2	4.01	63	418291	18.38	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.52%
20) 2-Butanone-d5	7.09	46	135530	27.43	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.86%
24) Chloroform-d	7.65	84	549684	22.54	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.16%
26) 1,2-Dichloroethane-d4	8.31	65	294392	21.65	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.60%
29) Benzene-d6	8.27	84	1074640	25.45	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.80%
33) 1,2-Dichloropropane-d6	9.27	67	351716	25.49	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.96%
37) Toluene-d8	10.33	98	854204	22.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.80%
38) trans-1,3-Dichloropropene-	10.58	79	100922	17.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.12%
39) 2-Hexanone-d5	10.93	63	92938	27.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	55.34%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	256127	17.12	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	68.48%
61) 1,2-Dichlorobenzene-d4	13.86	152	210653	25.29	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.16%

Target Compounds					Qvalue
16) Methylene chloride	4.91	84	59682	3.335 ug/L	95
47) Tetrachloroethene	10.87	164	37071	3.806 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW080218\
Data File : VW004446.D
Acq On : 03 Aug 2018 14:13
Operator : SY/AP
Sample : J4334-01
Misc : 5.45G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BEY64

Quant Time: Aug 04 02:12:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 04 01:26:48 2018
Response via : Initial Calibration

