

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080422\
 Data File : VW024162.D
 Acq On : 04 Aug 2022 15:24
 Operator : SY/VA
 Sample : VW0804SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK569

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/07/2022
 Supervised By :Mahesh Dadoda 08/08/2022

Quant Time: Aug 05 02:51:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 05 01:00:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.835	114	306221	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	298298	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	139829	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	156315	24.421	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	97.680%		
7) Chloroethane-d5	2.873	69	93165	23.365	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	93.480%		
11) 1,1-Dichloroethene-d2	4.001	63	121200	16.615	ug/L	-0.01
Spiked Amount 25.000	Range 45 - 110		Recovery =	66.440%		
21) 2-Butanone-d5	7.074	46	52199	43.834	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	87.660%		
24) Chloroform-d	7.641	84	196380	21.195	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	84.760%		
26) 1,2-Dichloroethane-d4	8.299	65	107814m	20.959	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	83.840%		
32) Benzene-d6	8.268	84	384090	21.833	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	87.320%		
36) 1,2-Dichloropropane-d6	9.268	67	113826	21.914	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	87.640%		
41) Toluene-d8	10.323	98	353106	21.474	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	85.880%		
43) trans-1,3-Dichloroprop...	10.573	79	46969	20.645	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	82.560%		
47) 2-Hexanone-d5	10.920	63	40611	42.329	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	84.660%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	103271	20.492	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	81.960%		
66) 1,2-Dichlorobenzene-d4	13.853	152	120389	23.528	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	94.120%		
Target Compounds					Qvalue	
13) Acetone	4.117	43	3861m	4.404	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080422\
Data File : VW024162.D
Acq On : 04 Aug 2022 15:24
Operator : SY/VA
Sample : VW0804SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK569

Quant Time: Aug 05 02:51:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 05 01:00:42 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/07/2022
Supervised By :Mahesh Dadoda 08/08/2022

