

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080525\
 Data File : VW032008.D
 Acq On : 05 Aug 2025 12:13
 Operator : SY/MD
 Sample : Q2760-01RE
 Misc : 5.23g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 NB-07-08042025RE

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/06/2025
 Supervised By :Semsettin Yesilyurt 08/06/2025

Quant Time: Aug 06 03:53:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 23 08:18:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	55750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	100720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	86486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	34475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	33929	40.231	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	80.460%		
35) Dibromofluoromethane	7.898	113	31258	45.672	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	91.340%		
50) Toluene-d8	10.325	98	102205	39.457	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	78.920%		
62) 4-Bromofluorobenzene	12.617	95	34332	34.518	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	69.040%		
Target Compounds					Qvalue	
16) Acetone	4.155	43	3836m	19.168	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080525\
Data File : VW032008.D
Acq On : 05 Aug 2025 12:13
Operator : SY/MD
Sample : Q2760-01RE
Misc : 5.23g/5mL/MSVOA_W/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
NB-07-08042025RE

Quant Time: Aug 06 03:53:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W072225S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 23 08:18:46 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/06/2025
Supervised By :Semsettin Yesilyurt 08/06/2025

