

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081418\
 Data File : VW004655.D
 Acq On : 13 Aug 2018 18:53
 Operator : SY/AP
 Sample : J4450-06
 Misc : 6.33G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAB05

Quant Time: Aug 14 02:26:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 14 01:20:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	953016	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	971867	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	466751	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	267721	22.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.36%
7) Chloroethane-d5	2.89	69	225457	22.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.44%
10) 1,1-Dichloroethene-d2	4.01	63	401084	16.66	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.64%
20) 2-Butanone-d5	7.09	46	290998	55.68	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.36%
24) Chloroform-d	7.65	84	597411	23.16	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.64%
26) 1,2-Dichloroethane-d4	8.31	65	362514	25.20	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.80%
29) Benzene-d6	8.28	84	1144463	22.14	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.56%
33) 1,2-Dichloropropane-d6	9.28	67	391545	23.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.68%
37) Toluene-d8	10.33	98	998174	21.90	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.60%
38) trans-1,3-Dichloropropene-	10.58	79	158400	22.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.20%
39) 2-Hexanone-d5	10.93	63	239773	58.29	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.58%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	474160	25.88	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.52%
61) 1,2-Dichlorobenzene-d4	13.86	152	436690	24.91	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.64%

Target Compounds						Ovalue
13) Acetone	4.14	43	202743	44.77	ug/L	99
21) 2-Butanone	7.18	43	53745	8.67	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081418\
Data File : VW004655.D
Acq On : 13 Aug 2018 18:53
Operator : SY/AP
Sample : J4450-06
Misc : 6.33G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAB05

Quant Time: Aug 14 02:26:53 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM080218S.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 14 01:20:10 2018
Response via : Initial Calibration

