

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081423\
 Data File : VW026729.D
 Acq On : 14 Aug 2023 12:29
 Operator : SY/MD
 Sample : 03968-25
 Misc : 4.50g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A48P2

Quant Time: Aug 14 23:26:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Aug 14 23:23:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	874997	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	808658	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	370312	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	204621	14.226	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.920%
7) Chloroethane-d5	2.893	69	172668	16.143	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.560%
11) 1,1-Dichloroethene-d2	4.027	65	123610	17.473	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.880%
21) 2-Butanone-d5	7.075	46	176503	42.693	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.380%
24) Chloroform-d	7.648	84	613092	20.877	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.520%
26) 1,2-Dichloroethane-d4	8.307	65	369522	22.801	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.200%
32) Benzene-d6	8.276	84	1244818	22.336	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.360%
36) 1,2-Dichloropropane-d6	9.276	67	407089	23.450	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.800%
41) Toluene-d8	10.325	98	1074217	22.454	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	10.575	79	156928	21.894	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.560%
47) 2-Hexanone-d5	10.922	63	86074	47.186	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	346354	25.263	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.040%
66) 1,2-Dichlorobenzene-d4	13.854	152	326016	24.597	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081423\
Data File : VW026729.D
Acq On : 14 Aug 2023 12:29
Operator : SY/MD
Sample : 03968-25
Misc : 4.50g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A48P2

Quant Time: Aug 14 23:26:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Aug 14 23:23:26 2023
Response via : Initial Calibration

