

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
 Data File : VW032093.D
 Acq On : 14 Aug 2025 15:01
 Operator : SY/MD
 Sample : Q2820-10
 Misc : 5.39g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SOIL-DUP-2

Quant Time: Aug 18 03:53:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	131941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	266004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.636	117	258211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	103366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	113358	59.405	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	118.800%
35) Dibromofluoromethane	7.904	113	97310	56.625	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	113.260%
50) Toluene-d8	10.325	98	318544	50.116	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	100.240%
62) 4-Bromofluorobenzene	12.617	95	102332	43.688	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	87.380%
Target Compounds						
						Qvalue
16) Acetone	4.167	43	21181	47.942	ug/l	97
17) Carbon Disulfide	4.430	76	14886	3.466	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
Data File : VW032093.D
Acq On : 14 Aug 2025 15:01
Operator : SY/MD
Sample : Q2820-10
Misc : 5.39g/5mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SOIL-DUP-2

Quant Time: Aug 18 03:53:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration

