

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
 Data File : VW032096.D
 Acq On : 14 Aug 2025 16:10
 Operator : SY/MD
 Sample : Q2820-13
 Misc : 9.13g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 10P-W

Quant Time: Aug 18 03:54:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	153502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	317351	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	347999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	179419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	120279	54.178	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	108.360%
35) Dibromofluoromethane	7.904	113	108290	52.819	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	105.640%
50) Toluene-d8	10.325	98	353584	46.628	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	93.260%
62) 4-Bromofluorobenzene	12.617	95	127022	45.455	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	90.900%
Target Compounds						
						Qvalue
16) Acetone	4.161	43	7948	15.463	ug/l	# 88
20) Methylene Chloride	4.960	84	13233	5.568	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
Data File : VW032096.D
Acq On : 14 Aug 2025 16:10
Operator : SY/MD
Sample : Q2820-13
Misc : 9.13g/5mL/MSVOA_W/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
10P-W

Quant Time: Aug 18 03:54:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration

