

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
 Data File : VW032098.D
 Acq On : 14 Aug 2025 16:55
 Operator : SY/MD
 Sample : Q2820-15
 Misc : 7.12g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 10P-S

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/19/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 18 03:55:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	68904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	131758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	126210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	77275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	50697	50.873	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery = 101.740%			
35) Dibromofluoromethane	7.898	113	46810	54.992	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery = 109.980%			
50) Toluene-d8	10.324	98	146293	46.466	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery = 92.940%			
62) 4-Bromofluorobenzene	12.617	95	64939	55.971	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery = 111.940%			
Target Compounds					Qvalue	
16) Acetone	4.161	43	3499m	15.165	ug/l	
20) Methylene Chloride	4.947	84	6349	5.952	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
Data File : VW032098.D
Acq On : 14 Aug 2025 16:55
Operator : SY/MD
Sample : Q2820-15
Misc : 7.12g/5mL/MSVOA_W/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
10P-S

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 08/19/2025
Supervised By : Mahesh Dadoda 08/19/2025

Quant Time: Aug 18 03:55:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration

