

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
 Data File : VW032081.D
 Acq On : 14 Aug 2025 10:00
 Operator : SY/MD
 Sample : VW0814SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0814SBL01

Quant Time: Aug 18 03:46:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	151415	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	296729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	295553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	134418	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	119802	54.707	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	109.420%
35) Dibromofluoromethane	7.898	113	104879	54.710	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	109.420%
50) Toluene-d8	10.325	98	349304	49.265	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.617	95	121836	46.629	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	93.260%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081425\
Data File : VW032081.D
Acq On : 14 Aug 2025 10:00
Operator : SY/MD
Sample : VW0814SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0814SBL01

Quant Time: Aug 18 03:46:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.959 min Scan# 996
Delta R.T. 0.000 min
Lab File: VW032081.D
Acq: 14 Aug 2025 10:00




Instrument :
MSVOA_W
ClientSampleId :
VW0814SBL01


Tgt Ion:168 Resp: 151415
Ion Ratio Lower Upper
168 100
99 65.4 56.0 84.0



#33
1,2-Dichloroethane-d4
Concen: 54.707 ug/l
RT: 8.313 min Scan# 1054
Delta R.T. -0.006 min
Lab File: VW032081.D
Acq: 14 Aug 2025 10:00

Tgt Ion: 65 Resp: 119802
Ion Ratio Lower Upper
65 100
67 50.0 0.0 102.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.849 min Scan# 1142

Delta R.T. -0.000 min

Lab File: VW032081.D

Acq: 14 Aug 2025 10:00

Instrument :

MSVOA_W

ClientSampleId :

VW0814SBL01





Tgt Ion:114 Resp: 296729

Ion Ratio Lower Upper

114 100

63 21.5 0.0 43.0

88 15.8 0.0 33.8



#35

Dibromofluoromethane

Concen: 54.710 ug/l

RT: 7.898 min Scan# 986

Delta R.T. -0.000 min

Lab File: VW032081.D

Acq: 14 Aug 2025 10:00




Tgt Ion:113 Resp: 104879


Ion Ratio Lower Upper

113 100

111 101.8 80.1 120.1

192 16.2 13.1 19.7









#50
Toluene-d8
Concen: 49.265 ug/l
RT: 10.325 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW032081.D
Acq: 14 Aug 2025 10:00




Instrument :
MSVOA_W
ClientSampleId :
VW0814SBL01


Tgt Ion: 98 Resp: 349304
Ion Ratio Lower Upper
98 100
100 64.3 51.1 76.7



#62
4-Bromofluorobenzene
Concen: 46.629 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. -0.000 min
Lab File: VW032081.D
Acq: 14 Aug 2025 10:00

Tgt Ion: 95 Resp: 121836
Ion Ratio Lower Upper
95 100
174 73.1 0.0 129.0
176 71.4 0.0 124.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.629 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW032081.D
Acq: 14 Aug 2025 10:00

Instrument :
MSVOA_W
ClientSampleId :
VW0814SBL01

Tgt Ion:117 Resp: 295553
Ion Ratio Lower Upper
117 100
82 52.0 45.0 67.4
119 27.5 24.2 36.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.556 min Scan# 1914
Delta R.T. -0.000 min
Lab File: VW032081.D
Acq: 14 Aug 2025 10:00

Tgt Ion:152 Resp: 134418
Ion Ratio Lower Upper
152 100
115 62.6 32.6 97.8
150 160.0 0.0 359.4

