

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081518\
 Data File : VW004676.D
 Acq On : 14 Aug 2018 19:16
 Operator : SY/AP
 Sample : VW0814SBL02
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 14 21:01:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081418S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 14 12:36:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	807393	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	818448	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	364102	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	318485	24.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.28%
7) Chloroethane-d5	2.89	69	238785	24.07	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.28%
10) 1,1-Dichloroethene-d2	4.01	63	430766	18.35	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.40%
20) 2-Butanone-d5	7.09	46	228645	65.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.68%
24) Chloroform-d	7.65	84	527119	24.36	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.44%
26) 1,2-Dichloroethane-d4	8.31	65	296025	25.82	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.28%
29) Benzene-d6	8.28	84	1069964	23.31	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.24%
33) 1,2-Dichloropropane-d6	9.27	67	337298	23.66	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.64%
37) Toluene-d8	10.33	98	931450	22.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.48%
38) trans-1,3-Dichloropropene-	10.58	79	132403	22.87	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.48%
39) 2-Hexanone-d5	10.93	63	168140	59.01	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.02%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	351698	26.30	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.20%
61) 1,2-Dichlorobenzene-d4	13.86	152	362417	26.24	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.96%

Target Compounds


					Ovalue
13) Acetone	4.14	43	7269	2.281 ug/L	94
16) Methylene chloride	4.91	84	35143	2.179 ug/L	97
49) 2-Hexanone	10.93	43	20572	3.227 ug/L #	67
59) 1,2,3-Trichloropropane	13.57	75	41674	4.108 ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081518\
Data File : VW004676.D
Acq On : 14 Aug 2018 19:16
Operator : SY/AP
Sample : VW0814SBL02
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 14 21:01:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081418S.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 14 12:36:48 2018
Response via : Initial Calibration





#13

Acetone

Concen: 2.281 ug/L

RT: 4.14 min Scan# 418

Delta R.T. 0.00 min

Lab File: VW004676.D

Acq: 14 Aug 2018 19:16

Instrument :

MSVOA_W





ClientSampled :

Tot Ion: 43 Resp: 7269

Ion Ratio Lower Upper

43 100

58 31.1 0.0 69.4



#16

Methylene chloride

Concen: 2.179 ug/L

RT: 4.91 min Scan# 545

Delta R.T. -0.01 min

Lab File: VW004676.D

Acq: 14 Aug 2018 19:16




Tgt Ion: 84 Resp: 35143


Ion Ratio Lower Upper

84 100

49 118.7 84.9 157.7

86 68.6 45.7 84.9









#49
2-Hexanone
Concen: 3.227 ug/L
RT: 10.93 min Scan# 1532
Delta R.T. -0.05 min
Lab File: VW004676.D
Acq: 14 Aug 2018 19:16

Instrument :
MSVOA_W
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
43	100		
58	32.7	44.4	66.6#
57	39.3	14.8	22.2#
100	4.1	9.4	14.2#



#59
1,2,3-Trichloropropane
Concen: 4.108 ug/L
RT: 13.57 min Scan# 1965
Delta R.T. 0.79 min
Lab File: VW004676.D
Acq: 14 Aug 2018 19:16

Tgt Ion	Ratio	Lower	Upper
75	100		
77	36.9	25.8	38.6

