

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081622\
 Data File : VW024258.D
 Acq On : 16 Aug 2022 17:02
 Operator : SY/VA
 Sample : N4221-03
 Misc : 6.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0KP4

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/17/2022
 Supervised By :Semsettin Yesilyurt 08/17/2022

Quant Time: Aug 17 01:03:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 17 01:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	342962	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	317251	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	130539	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	144138	20.106	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	80.440%		
7) Chloroethane-d5	2.873	69	94409	21.141	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	84.560%		
11) 1,1-Dichloroethene-d2	4.007	63	123492	15.115	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	60.480%		
21) 2-Butanone-d5	7.080	46	57089	42.805	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	85.600%		
24) Chloroform-d	7.647	84	215276	20.745	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	83.000%		
26) 1,2-Dichloroethane-d4	8.305	65	120443	20.905	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	83.640%		
32) Benzene-d6	8.269	84	396990	21.218	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	84.880%		
36) 1,2-Dichloropropane-d6	9.274	67	124493	22.536	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	90.160%		
41) Toluene-d8	10.323	98	358872	20.521	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	82.080%		
43) trans-1,3-Dichloroprop...	10.579	79	48593	20.083	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	80.320%		
47) 2-Hexanone-d5	10.921	63	42982	42.124	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	84.240%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	109348	20.402	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	81.600%		
66) 1,2-Dichlorobenzene-d4	13.853	152	108086	22.627	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	90.520%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	4.032	96	6858	1.663	ug/L #	1
13) Acetone	4.117	43	9659	9.837	ug/L	92
14) Carbon disulfide	4.373	76	17163	1.301	ug/L	95
19) 1,1-Dichloroethane	6.214	63	8223m	0.980	ug/L	
20) cis-1,2-Dichloroethene	7.165	96	12370	2.366	ug/L #	92
34) Trichloroethene	9.092	95	533662	107.754	ug/L	94
46) Tetrachloroethene	10.866	164	260939	65.080	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081622\
Data File : VW024258.D
Acq On : 16 Aug 2022 17:02
Operator : SY/VA
Sample : N4221-03
Misc : 6.10g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0KP4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/17/2022
Supervised By :Semsettin Yesilyurt 08/17/2022

Quant Time: Aug 17 01:03:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 17 01:00:38 2022
Response via : Initial Calibration

