

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082024\
 Data File : VW029961.D
 Acq On : 20 Aug 2024 16:08
 Operator : SY/MD
 Sample : P3626-19
 Misc : 4.68g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCYB0

Quant Time: Aug 21 00:47:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 21 00:44:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	281590	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	223179	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	58528	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	63211	15.812	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.240%
7) Chloroethane-d5	2.899	69	56451	19.779	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.120%
11) 1,1-Dichloroethene-d2	4.021	65	27470	15.053	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.200%
21) 2-Butanone-d5	7.075	46	32329	29.506	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	59.020%
24) Chloroform-d	7.648	84	140923	19.333	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.320%
26) 1,2-Dichloroethane-d4	8.307	65	79222	19.081	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.320%
32) Benzene-d6	8.276	84	247297	20.508	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.040%
36) 1,2-Dichloropropane-d6	9.270	67	81772	21.237	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.960%
41) Toluene-d8	10.319	98	204895	19.109	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.440%
43) trans-1,3-Dichloroprop...	10.575	79	28786	18.475	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.920%
47) 2-Hexanone-d5	10.922	63	23805	37.494	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	54687	19.128	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	76.520%
66) 1,2-Dichlorobenzene-d4	13.848	152	36766	20.231	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.920%
Target Compounds					Qvalue	
13) Acetone	4.125	43	25390	18.646	ug/L	93
42) Toluene	10.386	91	6386	0.409	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082024\
Data File : VW029961.D
Acq On : 20 Aug 2024 16:08
Operator : SY/MD
Sample : P3626-19
Misc : 4.68g/10mL/MSVOA_W/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCYB0

Quant Time: Aug 21 00:47:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 21 00:44:34 2024
Response via : Initial Calibration

