

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082024\
 Data File : VW029966.D
 Acq On : 20 Aug 2024 18:08
 Operator : SY/MD
 Sample : P3675-02
 Misc : 3.93g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCXY1

Quant Time: Aug 21 00:48:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 21 00:44:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	239257	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	181018	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	53725	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	62219	18.318	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.280%
7) Chloroethane-d5	2.905	69	55830	23.023	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.080%
11) 1,1-Dichloroethene-d2	4.021	65	26704	17.223	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.880%
21) 2-Butanone-d5	7.087	46	35631	38.273	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.540%
24) Chloroform-d	7.648	84	134027	21.640	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.560%
26) 1,2-Dichloroethane-d4	8.307	65	78033	22.120	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.480%
32) Benzene-d6	8.276	84	236165	24.146	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.600%
36) 1,2-Dichloropropane-d6	9.270	67	76842	24.605	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.400%
41) Toluene-d8	10.325	98	188079	21.627	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.520%
43) trans-1,3-Dichloroprop...	10.575	79	27926	22.098	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.400%
47) 2-Hexanone-d5	10.922	63	26363	51.194	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	57272	24.698	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.800%
66) 1,2-Dichlorobenzene-d4	13.848	152	32522	19.495	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.000%
Target Compounds						
13) Acetone	4.143	43	90115	77.889	ug/L	96
15) Methyl Acetate	4.685	43	6653	3.452	ug/L #	86
22) 2-Butanone	7.191	43	6038m	4.073	ug/L	
42) Toluene	10.386	91	18235	1.440	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082024\
Data File : VW029966.D
Acq On : 20 Aug 2024 18:08
Operator : SY/MD
Sample : P3675-02
Misc : 3.93g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCXY1

Quant Time: Aug 21 00:48:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 21 00:44:34 2024
Response via : Initial Calibration

