

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082224\
 Data File : VW029979.D
 Acq On : 22 Aug 2024 11:30
 Operator : SY/MD
 Sample : VIBLK509
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK509

Quant Time: Aug 23 01:49:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 23 01:48:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	231455	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	230157	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	100069	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	61489	18.713	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	74.840%		
7) Chloroethane-d5	2.905	69	53983	23.011	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	92.040%		
11) 1,1-Dichloroethene-d2	4.021	65	29541	19.695	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	78.760%		
21) 2-Butanone-d5	7.087	46	34744	38.578	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	77.160%		
24) Chloroform-d	7.648	84	153647	25.644	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	102.560%		
26) 1,2-Dichloroethane-d4	8.307	65	85919	25.176	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	100.720%		
32) Benzene-d6	8.276	84	276197	22.210	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	88.840%		
36) 1,2-Dichloropropane-d6	9.270	67	88278	22.232	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	88.920%		
41) Toluene-d8	10.325	98	243851	22.053	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	88.200%		
43) trans-1,3-Dichloroprop...	10.575	79	33911	21.105	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	84.400%		
47) 2-Hexanone-d5	10.922	63	26163	39.958	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	79.920%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	76720	26.021	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	104.080%		
66) 1,2-Dichlorobenzene-d4	13.848	152	85686	27.576	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	110.320%		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	13427	2.775	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082224\
Data File : VW029979.D
Acq On : 22 Aug 2024 11:30
Operator : SY/MD
Sample : VIBLK509
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK509

Quant Time: Aug 23 01:49:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 23 01:48:12 2024
Response via : Initial Calibration

