

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
 Data File : VW032121.D
 Acq On : 22 Aug 2025 19:02
 Operator : SY/MD
 Sample : Q2941-01
 Misc : 6.00g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SOIL-COMP(1-2)

Quant Time: Aug 25 04:01:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	179471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	349264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	349840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	177790	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	133163	51.302	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	= 102.600%		
35) Dibromofluoromethane	7.904	113	119487	52.955	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	= 105.920%		
50) Toluene-d8	10.324	98	422427	50.616	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	= 101.240%		
62) 4-Bromofluorobenzene	12.617	95	169750	55.194	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	= 110.380%		
Target Compounds						
52) Toluene	10.391	92	66925	10.746	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
Data File : VW032121.D
Acq On : 22 Aug 2025 19:02
Operator : SY/MD
Sample : Q2941-01
Misc : 6.00g/5mL/MSVOA_W/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SOIL-COMP(1-2)

Quant Time: Aug 25 04:01:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration

