

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
 Data File : VW032123.D
 Acq On : 22 Aug 2025 19:46
 Operator : SY/MD
 Sample : Q2928-02
 Misc : 7.16g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 27-BELL

Quant Time: Aug 25 04:02:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	183461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	365151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	344552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	168107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	131254	49.467	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	98.940%		
35) Dibromofluoromethane	7.904	113	117096	49.638	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	99.280%		
50) Toluene-d8	10.325	98	418031	47.910	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	95.820%		
62) 4-Bromofluorobenzene	12.617	95	157703	49.046	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	98.100%		
Target Compounds						
20) Methylene Chloride	4.960	84	18388	6.474	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
Data File : VW032123.D
Acq On : 22 Aug 2025 19:46
Operator : SY/MD
Sample : Q2928-02
Misc : 7.16g/5mL/MSVOA_W/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
27-BELL

Quant Time: Aug 25 04:02:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration

