

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
 Data File : VW032115.D
 Acq On : 22 Aug 2025 16:50
 Operator : SY/MD
 Sample : Q2933-03
 Misc : 5.49g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 29-SEPTIC

Quant Time: Aug 25 03:59:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 04:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	153557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	304549	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	300129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	160095	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	125999	56.734	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	113.460%
35) Dibromofluoromethane	7.904	113	111334	56.586	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	113.180%
50) Toluene-d8	10.325	98	372831	51.233	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	102.460%
62) 4-Bromofluorobenzene	12.617	95	133664	49.842	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	99.680%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
Data File : VW032115.D
Acq On : 22 Aug 2025 16:50
Operator : SY/MD
Sample : Q2933-03
Misc : 5.49g/5mL/MSVOA_W/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
29-SEPTIC

Quant Time: Aug 25 03:59:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 04:04:48 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.959 min Scan# 996
Delta R.T. 0.000 min
Lab File: VW032115.D
Acq: 22 Aug 2025 16:50




Instrument :
MSVOA_W
ClientSampleId :
29-SEPTIC


Tgt Ion:168 Resp: 153557
Ion Ratio Lower Upper
168 100
99 63.5 56.0 84.0



#33
1,2-Dichloroethane-d4
Concen: 56.734 ug/l
RT: 8.319 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VW032115.D
Acq: 22 Aug 2025 16:50

Tgt Ion: 65 Resp: 125999
Ion Ratio Lower Upper
65 100
67 49.5 0.0 102.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.849 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VW032115.D
Acq: 22 Aug 2025 16:50




Instrument :
MSVOA_W
ClientSampleId :
29-SEPTIC


Tgt Ion:114 Resp: 304549
Ion Ratio Lower Upper
114 100
63 20.7 0.0 43.0
88 15.9 0.0 33.8



#35
Dibromofluoromethane
Concen: 56.586 ug/l
RT: 7.904 min Scan# 987
Delta R.T. 0.006 min
Lab File: VW032115.D
Acq: 22 Aug 2025 16:50

Tgt Ion:113 Resp: 111334
Ion Ratio Lower Upper
113 100
111 103.1 80.1 120.1
192 18.1 13.1 19.7









#50
Toluene-d8
Concen: 51.233 ug/l
RT: 10.325 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW032115.D
Acq: 22 Aug 2025 16:50




Instrument :
MSVOA_W
ClientSampleId :
29-SEPTIC


Tgt Ion: 98 Resp: 372831
Ion Ratio Lower Upper
98 100
100 65.3 51.1 76.7



#62
4-Bromofluorobenzene
Concen: 49.842 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. -0.000 min
Lab File: VW032115.D
Acq: 22 Aug 2025 16:50

Tgt Ion: 95 Resp: 133664
Ion Ratio Lower Upper
95 100
174 80.1 0.0 129.0
176 75.4 0.0 124.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.635 min Scan# 11

Delta R.T. 0.006 min

Lab File: VW032115.D

Acq: 22 Aug 2025 16:50

Instrument :

MSVOA_W

ClientSampleId :

29-SEPTIC





Tgt Ion:117 Resp: 300129

Ion Ratio Lower Upper

117 100

82 52.6 45.0 67.4

119 32.4 24.2 36.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.556 min Scan# 1914

Delta R.T. -0.000 min

Lab File: VW032115.D

Acq: 22 Aug 2025 16:50

Tgt Ion:152 Resp: 160095

Ion Ratio Lower Upper

152 100

115 62.0 32.6 97.8

150 153.3 0.0 359.4

