



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082225\
Data File : VW032106.D
Acq On : 22 Aug 2025 10:45
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Title : SW846 8260
Last Update : Tue Aug 12 04:04:48 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1750

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	20861	PASS
75	95	30	60	56.7	53779	PASS
95	95	100	100	100.0	94904	PASS
96	95	5	9	7.2	6877	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.7	68000	PASS
175	174	5	9	8.3	5635	PASS
176	174	95	101	99.9	67899	PASS
177	176	5	9	6.5	4440	PASS