

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082324\
 Data File : VW030002.D
 Acq On : 23 Aug 2024 11:55
 Operator : SY/MD
 Sample : P3673-20
 Misc : 5.81g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCXZ4

Quant Time: Aug 24 05:01:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Aug 24 05:00:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	403880	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	381412	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	159858	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	71471	12.465	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	49.840%
7) Chloroethane-d5	2.899	69	62977	15.384	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.520%
11) 1,1-Dichloroethene-d2	4.027	65	33788	12.909	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.640%
21) 2-Butanone-d5	7.087	46	42494	27.040	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.080%
24) Chloroform-d	7.648	84	185389	17.732	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	70.920%
26) 1,2-Dichloroethane-d4	8.307	65	102385	17.193	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	68.760%#
32) Benzene-d6	8.276	84	331116	16.067	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.280%
36) 1,2-Dichloropropane-d6	9.276	67	107493	16.335	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	65.360%#
41) Toluene-d8	10.324	98	301444	16.451	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	65.800%
43) trans-1,3-Dichloroprop...	10.581	79	42139	15.826	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.320%
47) 2-Hexanone-d5	10.922	63	34570	31.860	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	91717	18.771	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.080%
66) 1,2-Dichlorobenzene-d4	13.848	152	96704	19.482	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	77.920%
Target Compounds						
13) Acetone	4.149	43	15230	7.798	ug/L	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082324\
Data File : VW030002.D
Acq On : 23 Aug 2024 11:55
Operator : SY/MD
Sample : P3673-20
Misc : 5.81g/10mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCXZ4

Quant Time: Aug 24 05:01:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Aug 24 05:00:17 2024
Response via : Initial Calibration

