

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082619\
 Data File : VW012182.D
 Acq On : 26 Aug 2019 20:23
 Operator : SY/VA
 Sample : K4475-02
 Misc : 6.11G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-2-E-(0)

Quant Time: Aug 27 05:56:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	211645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	355501	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	288904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	104316	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	139495	55.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.88%	
35) Dibromofluoromethane	7.88	113	111717	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
50) Toluene-d8	10.32	98	435994	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
62) 4-Bromofluorobenzene	12.62	95	128262	42.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.58%	
Target Compounds						
16) Acetone	4.14	43	31796	49.440	ug/l	99
20) Methylene Chloride	4.92	84	21919	9.015	ug/l	95
29) Tetrahydrofuran	7.54	42	5122	9.769	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082619\
Data File : VW012182.D
Acq On : 26 Aug 2019 20:23
Operator : SY/VA
Sample : K4475-02
Misc : 6.11G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T1-2-E-(0)

Quant Time: Aug 27 05:56:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

