

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082622\
 Data File : VW024357.D
 Acq On : 26 Aug 2022 18:16
 Operator : SY/VA
 Sample : N4344-13
 Misc : 5.28g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1




Instrument :
 MSVOA_W
 ClientSampleId :
 DBVR2

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/30/2022

Quant Time: Aug 27 00:42:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Aug 27 00:35:21 2022
 Response via : Initial Calibration

08/30/2022
 Supervised By :Mahesh
 Dadoda
 08/30/2022



TIC: VW024357.D\data.ms

(13) Acetone (T)

4.123min (+ 0.006) 3.53 ug/L

response 902

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	29.20	39.36
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082622\
 Data File : VW024357.D
 Acq On : 26 Aug 2022 18:16
 Operator : SY/VA
 Sample : N4344-13
 Misc : 5.28g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1




Instrument :
 MSVOA_W
 ClientSampleId :
 DBVR2

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/30/2022

Quant Time: Aug 27 00:42:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Aug 27 00:35:21 2022
 Response via : Initial Calibration

08/30/2022
 Supervised By :Mahesh
 Dadoda
 08/30/2022



TIC: VW024357.D\data.ms

(13) Acetone (T)

4.123min (+ 0.006) 8.71 ug/L m

response 2226

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	29.20	15.95
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082622\
 Data File : VW024357.D
 Acq On : 26 Aug 2022 18:16
 Operator : SY/VA
 Sample : N4344-13
 Misc : 5.28g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBVR2

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/30/2022

Quant Time: Aug 27 00:42:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Aug 27 00:35:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----08/30/2022						
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	84341	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	95714	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	39618	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	65736	28.274	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	113.080%	
7) Chloroethane-d5	2.867	69	32192	18.480	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	73.920%	
11) 1,1-Dichloroethene-d2	4.001	63	35349	17.531	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	70.120%	
21) 2-Butanone-d5	7.080	46	18281	42.690	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	85.380%	
24) Chloroform-d	7.647	84	55632	20.517	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	82.080%	
26) 1,2-Dichloroethane-d4	8.305	65	35402	21.530	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	86.120%	
32) Benzene-d6	8.275	84	102747	18.736	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	74.960%	
36) 1,2-Dichloropropane-d6	9.274	67	31819	18.657	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	74.640%	
41) Toluene-d8	10.323	98	87709	16.822	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	67.280%	
43) trans-1,3-Dichloroprop...	10.579	79	12729	15.949	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	63.800%	
47) 2-Hexanone-d5	10.921	63	11461	30.030	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	60.060%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	33739	17.473	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	69.880%	
66) 1,2-Dichlorobenzene-d4	13.853	152	29738	19.933	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	79.720%	
Target Compounds						
13) Acetone	4.123	43	2226m	8.711	ug/L	
16) Methylene chloride	4.909	84	8117	5.738	ug/L	82
46) Tetrachloroethene	10.860	164	22546	18.483	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082622\
Data File : VW024357.D
Acq On : 26 Aug 2022 18:16
Operator : SY/VA
Sample : N4344-13
Misc : 5.28g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBVR2

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
Supervised By :Mahesh Dadoda 08/30/2022

08/30/2022
Supervised By :Mahesh
Dadoda

08/30/2022

Quant Time: Aug 27 00:42:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Aug 27 00:35:21 2022
Response via : Initial Calibration

