

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082624\
 Data File : VW030027.D
 Acq On : 26 Aug 2024 14:30
 Operator : SY/MD
 Sample : VIBLK513
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK513

Quant Time: Aug 27 05:36:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Aug 27 05:31:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	316959	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	290860	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	137580	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	117729	24.960	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.840%
7) Chloroethane-d5	2.899	69	83639	24.984	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.920%
11) 1,1-Dichloroethene-d2	4.021	65	49764	24.694	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	98.760%
21) 2-Butanone-d5	7.088	46	51551	54.827	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.660%
24) Chloroform-d	7.648	84	211603	24.756	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.040%
26) 1,2-Dichloroethane-d4	8.301	65	116471	26.836	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.360%
32) Benzene-d6	8.276	84	438797	25.558	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.240%
36) 1,2-Dichloropropane-d6	9.270	67	128721	25.590	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.360%
41) Toluene-d8	10.319	98	401911	25.720	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.880%
43) trans-1,3-Dichloroprop...	10.575	79	54300	26.243	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	104.960%
47) 2-Hexanone-d5	10.922	63	45006	58.461	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.920%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	101122	27.305	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.200%
66) 1,2-Dichlorobenzene-d4	13.849	152	130757	26.276	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.120%
Target Compounds						
					Qvalue	
16) Methylene chloride	4.923	84	29407	5.616	ug/L	95
70) 1,2,4-trichlorobenzene	15.129	180	6823	1.290	ug/L	98
72) 1,2,3-Trichlorobenzene	15.550	180	8224	1.772	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082624\
Data File : VW030027.D
Acq On : 26 Aug 2024 14:30
Operator : SY/MD
Sample : VIBLK513
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK513

Quant Time: Aug 27 05:36:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Aug 27 05:31:07 2024
Response via : Initial Calibration

