

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082824\
 Data File : VW030071.D
 Acq On : 28 Aug 2024 13:09
 Operator : SY/MD
 Sample : P3721-13
 Misc : 3.67g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 29 01:14:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Aug 29 01:10:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/L	-8.84
28) Chlorobenzene-d5	0.000	117	0	0.000	ug/L	-11.63
58) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/L	-13.56
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.000%#
7) Chloroethane-d5	0.000	69	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.000%#
11) 1,1-Dichloroethene-d2	0.000	65	0	0.000	ug/L	
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.000%#
21) 2-Butanone-d5	0.000	46	0	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
24) Chloroform-d	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	40 - 150	Recovery	=	0.000%#
26) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/L	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.000%#
32) Benzene-d6	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	20 - 135	Recovery	=	0.000%#
36) 1,2-Dichloropropane-d6	0.000	67	0	0.000	ug/L	
Spiked Amount	25.000	Range	70 - 120	Recovery	=	0.000%#
41) Toluene-d8	0.000	98	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 130	Recovery	=	0.000%#
43) trans-1,3-Dichloroprop...	0.000	79	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.000%#
47) 2-Hexanone-d5	0.000	63	0	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
56) 1,1,2,2-Tetrachloroeth...	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	45 - 120	Recovery	=	0.000%#
66) 1,2-Dichlorobenzene-d4	0.000	152	0	0.000	ug/L	
Spiked Amount	25.000	Range	75 - 120	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082824\
Data File : VW030071.D
Acq On : 28 Aug 2024 13:09
Operator : SY/MD
Sample : P3721-13
Misc : 3.67g/10mL/MSVOA_W/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 29 01:14:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Aug 29 01:10:06 2024
Response via : Initial Calibration

