

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082824\
 Data File : VW030077.D
 Acq On : 28 Aug 2024 15:33
 Operator : SY/MD
 Sample : VIBLK515
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK515

Quant Time: Aug 29 01:16:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Aug 29 01:10:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	180441	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	167527	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	77609	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	59940	22.322	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.280%
7) Chloroethane-d5	2.899	69	45112	23.671	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.680%
11) 1,1-Dichloroethene-d2	4.021	65	28148	24.536	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	98.160%
21) 2-Butanone-d5	7.075	46	32562	60.833	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.660%
24) Chloroform-d	7.648	84	121402	24.949	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	8.300	65	66972	27.105	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.440%
32) Benzene-d6	8.276	84	242609	24.534	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.120%
36) 1,2-Dichloropropane-d6	9.276	67	76772	26.499	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.000%
41) Toluene-d8	10.318	98	217256	24.138	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.560%
43) trans-1,3-Dichloroprop...	10.574	79	30022	25.191	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	100.760%
47) 2-Hexanone-d5	10.922	63	25432	57.356	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	61727	28.938	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	115.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	71657	25.527	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.120%
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	7811	2.620	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082824\
Data File : VW030077.D
Acq On : 28 Aug 2024 15:33
Operator : SY/MD
Sample : VIBLK515
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK515

Quant Time: Aug 29 01:16:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Aug 29 01:10:06 2024
Response via : Initial Calibration

