

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082825\
 Data File : VW032164.D
 Acq On : 28 Aug 2025 16:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 29 01:22:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.330	0.289	12.4	91	0.00
3 P	Chloromethane	0.346	0.292	15.6	84	0.00
4 C	Vinyl Chloride	0.450	0.388	13.8#	86	0.00
5 T	Bromomethane	0.401	0.342	14.7	90	0.00
6 T	Chloroethane	0.317	0.283	10.7	88	0.00
7 T	Trichlorofluoromethane	0.445	0.372	16.4	76	0.00
8 T	Diethyl Ether	0.332	0.303	8.7	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.449	10.6	87	0.00
10 T	Methyl Iodide	0.764	0.706	7.6	92	0.00
11 T	Tert butyl alcohol	0.044	0.041	6.8	91	0.00
12 CM	1,1-Dichloroethene	0.528	0.481	8.9#	88	0.00
13 T	Acrolein	0.067	0.050	25.4#	74	0.00
14 T	Allyl chloride	0.751	0.681	9.3	87	0.00
15 T	Acrylonitrile	0.162	0.161	0.6	97	0.00
16 T	Acetone	0.155	0.130	16.1	81	0.00
17 T	Carbon Disulfide	1.322	1.191	9.9	86	0.00
18 T	Methyl Acetate	0.493	0.537	-8.9	102	0.00
19 T	Methyl tert-butyl Ether	0.998	0.966	3.2	90	0.00
20 T	Methylene Chloride	0.725	0.593	18.2	91	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.570	3.4	92	0.00
22 T	Diisopropyl ether	1.634	1.615	1.2	92	0.00
23 T	Vinyl Acetate	1.111	1.080	2.8	90	0.00
24 P	1,1-Dichloroethane	1.070	1.009	5.7	90	0.00
25 T	2-Butanone	0.219	0.220	-0.5	95	0.00
26 T	2,2-Dichloropropane	0.566	0.510	9.9	82	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.689	4.0	91	0.00
28 T	Bromochloromethane	0.490	0.489	0.2	99	0.00
29 T	Tetrahydrofuran	0.139	0.143	-2.9	99	0.00
30 C	Chloroform	1.221	1.183	3.1#	93	0.00
31 T	Cyclohexane	0.878	0.769	12.4	87	0.00
32 T	1,1,1-Trichloroethane	0.891	0.830	6.8	89	0.00
33 S	1,2-Dichloroethane-d4	0.715	0.692	3.2	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.351	0.339	3.4	96	0.00
36 T	1,1-Dichloropropene	0.455	0.445	2.2	89	0.00
37 T	Ethyl Acetate	0.273	0.276	-1.1	95	0.00
38 T	Carbon Tetrachloride	0.488	0.461	5.5	86	0.00
39 T	Methylcyclohexane	0.539	0.506	6.1	84	0.00
40 TM	Benzene	1.406	1.380	1.8	92	0.00
41 T	Methacrylonitrile	0.152	0.149	2.0	96	0.00
42 TM	1,2-Dichloroethane	0.487	0.486	0.2	95	0.00
43 T	Isopropyl Acetate	0.512	0.509	0.6	92	0.00
44 TM	Trichloroethene	0.354	0.337	4.8	92	0.00
45 C	1,2-Dichloropropane	0.339	0.337	0.6#	93	0.00
46 T	Dibromomethane	0.238	0.233	2.1	95	0.00
47 T	Bromodichloromethane	0.540	0.534	1.1	92	0.00
48 T	Methyl methacrylate	0.236	0.243	-3.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082825\
 Data File : VW032164.D
 Acq On : 28 Aug 2025 16:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 29 01:22:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	104	0.00
50 S	Toluene-d8	1.231	1.196	2.8	92	0.00
51 T	4-Methyl-2-Pentanone	0.289	0.306	-5.9	98	0.00
52 CM	Toluene	0.919	0.907	1.3#	92	0.00
53 T	t-1,3-Dichloropropene	0.477	0.469	1.7	89	0.00
54 T	cis-1,3-Dichloropropene	0.544	0.537	1.3	89	0.00
55 T	1,1,2-Trichloroethane	0.309	0.301	2.6	95	0.00
56 T	Ethyl methacrylate	0.395	0.416	-5.3	94	0.00
57 T	1,3-Dichloropropane	0.518	0.506	2.3	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.211	0.210	0.5	96	0.00
59 T	2-Hexanone	0.201	0.218	-8.5	99	0.00
60 T	Dibromochloromethane	0.374	0.369	1.3	91	0.00
61 T	1,2-Dibromoethane	0.304	0.297	2.3	91	0.00
62 S	4-Bromofluorobenzene	0.455	0.439	3.5	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.329	0.304	7.6	90	0.00
65 PM	Chlorobenzene	1.138	1.056	7.2	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.355	5.3	92	0.00
67 C	Ethyl Benzene	1.865	1.801	3.4#	93	0.00
68 T	m/p-Xylenes	0.718	0.695	3.2	91	0.00
69 T	o-Xylene	0.679	0.680	-0.1	94	0.00
70 T	Styrene	1.197	1.155	3.5	91	0.00
71 P	Bromoform	0.223	0.221	0.9	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.397	3.183	6.3	87	0.00
74 T	N-amyl acetate	0.986	0.958	2.8	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.807	0.796	1.4	97	0.00
76 T	1,2,3-Trichloropropane	0.626	0.585	6.5	96	0.00
77 T	Bromobenzene	0.866	0.844	2.5	93	0.00
78 T	n-propylbenzene	4.124	4.031	2.3	92	0.00
79 T	2-Chlorotoluene	2.522	2.436	3.4	92	0.00
80 T	1,3,5-Trimethylbenzene	2.876	2.780	3.3	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.233	0.224	3.9	90	0.00
82 T	4-Chlorotoluene	2.699	2.577	4.5	90	0.00
83 T	tert-Butylbenzene	2.439	2.424	0.6	89	0.00
84 T	1,2,4-Trimethylbenzene	2.975	2.980	-0.2	93	0.00
85 T	sec-Butylbenzene	3.641	3.483	4.3	88	0.00
86 T	p-Isopropyltoluene	3.096	3.037	1.9	90	0.00
87 T	1,3-Dichlorobenzene	1.744	1.670	4.2	92	0.00
88 T	1,4-Dichlorobenzene	1.748	1.612	7.8	92	0.00
89 T	n-Butylbenzene	2.952	2.873	2.7	90	0.00
90 T	Hexachloroethane	0.566	0.541	4.4	92	0.00
91 T	1,2-Dichlorobenzene	1.544	1.484	3.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.143	0.142	0.7	101	0.00
93 T	1,2,4-Trichlorobenzene	0.987	0.941	4.7	90	0.00
94 T	Hexachlorobutadiene	0.489	0.448	8.4	84	0.00
95 T	Naphthalene	2.233	2.211	1.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082825\
Data File : VW032164.D
Acq On : 28 Aug 2025 16:36
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Aug 29 01:22:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.918	0.871	5.1	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6