




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024488.D
Acq On : 30 Aug 2022 11:42
Operator : SY/VA
Sample : N4346-11RE
Misc : 6.27g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 01:20:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Aug 30 07:05:43 2022
Response via : Initial Calibration



TIC: VW024488.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

4.007min (-0.006) 16.06 ug/L m

response 154116

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	70.50	0.37#
65.00	22.70	0.09#
0.00	0.00	0.00