




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024504.D
Acq On : 30 Aug 2022 20:53
Operator : SY/VA
Sample : N4345-05RE
Misc : 4.29g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 02:57:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 31 02:55:40 2022
Response via : Initial Calibration



TIC: VW024504.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.421min (+ 0.115) 0.05 ug/L

response 367

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	51.10	52.04
51.00	96.50	121.53
0.00	0.00	0.00